

TOXL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005946
 12/20/2019 12:14:33

Auto Calibration

pg 2 of 2

<<<<< 3um >>>>>			<<<<< 9um >>>>>		
-----			-----		
Zero Order Coef	-391.96			-349.15	
First Order Coef	2860.61			1392.11	
Second Order Coef	32.46			13.72	
-----			-----		
Act	Fit	Residual	Act	Fit	Residual
(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)
0.000	-0.001	0.0010	0.000	-0.001	0.0007
0.040	0.041	-0.0012	0.040	0.041	-0.0011
0.081	0.082	-0.0007	0.081	0.081	-0.0001
0.151	0.150	0.0012	0.151	0.150	0.0006
0.301	0.301	-0.0002	0.301	0.301	-0.0001
-----			-----		

<<<<< 3um >>>>>			<<<<< 9um >>>>>		
-----			-----		
Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1					
Sample					
Sample #1		4226.00		4342.00	
Sample #2		4041.00		4167.00	
Sample #3		4086.00		4184.00	
Sample #4		4186.00		4236.00	
Avg		4104.3335		4195.6665	
STD DEV		74.2181		35.9490	
REL STD DEV		1.808		0.857	
H2O adjust (mg/l*10k)		657		566	

Atmospheric Pressure = 957

*****CALIBRATION SUCCESSFUL*****

Charles E. Edr
 12/20/19

TOXL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005946
 12/20/2019 12:14:33

Auto Calibration
 Max Power Res Value = 18
 Auto Range Res Value = 5

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.14.00 09/16
01/16/2020 10:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:17
02 Std. Sol.	0.017	10:18
03 Room Air	0.000	10:18
04 Std. Sol.	0.017	10:19
05 Room Air	0.000	10:20
06 Std. Sol.	0.017	10:20
07 Room Air	0.000	10:21

08 Sim Temp = 34.0°C

Simul Ser No = MP3071
Std Sol No = ~~2108~~10D
County = 08 ^{CEE} 2018 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.14.00 09/16
01/16/2020 10:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:34
02 Std. Sol.	0.249	10:35
03 Room Air	0.000	10:36
04 Std. Sol.	0.250	10:36
05 Room Air	0.000	10:37
06 Std. Sol.	0.248	10:38
07 Room Air	0.000	10:38

08 Sim Temp = 34.0°C

Simul Ser No = DR7351
Std Sol No = 201803G
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.250 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.14.00 09/16
01/16/2020 12:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:05
02 Std. Gas	0.080	12:05
03 Room Air	0.000	12:06
04 Std. Gas	0.081	12:06
05 Room Air	0.000	12:07
06 Std. Gas	0.080	12:07
07 Room Air	0.000	12:08

Lot No = 34917080A3
Cyl No = 2
Exp Date = 02/05/2020
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.14.00 09/16
01/16/2020 12:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:08
02 Std. Gas	0.080	12:09
03 Room Air	0.000	12:09
04 Std. Gas	0.080	12:10
05 Room Air	0.000	12:10
06 Std. Gas	0.080	12:11
07 Room Air	0.000	12:11

Lot No = 34917080A3
Cyl No = 2
Exp Date = 02/05/2020
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.14.00 09/16
01/16/2020 12:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:12
02 Std. Gas	0.080	12:12
03 Room Air	0.000	12:13
04 Std. Gas	0.079	12:13
05 Room Air	0.000	12:14
06 Std. Gas	0.080	12:14
07 Room Air	0.000	12:15

Lot No = 34917080A3
Cyl No = 2
Exp Date = 02/05/2020
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000