



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Sand Lemieux</i>		
Location <i>Burlington Norton Detention Center</i>	Intoxilyzer® Serial Number <i>80-005952</i>	
Gas Lot Number <i>24119080A1</i>	Gas Cylinder Number <i>29</i>	Gas Expiration Date <i>11/5/2021</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.082* AC
 - B. *0.082* AC
 - C. *0.083* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i> 4824	Date <i>1/22/2020</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>3/4/2020</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BMDC 8164.14.00 09/16
01/22/2020 16:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:39
02 Std. Gas	0.082	16:39
03 Room Air	0.000	16:40
04 Std. Gas	0.082	16:40
05 Room Air	0.000	16:41
06 Std. Gas	0.083	16:41
07 Room Air	0.000	16:42

Lot No = 24119080A1
Cyl No = 29
Exp Date = 11/05/2021
County = 08 Oper No. = 130315


Operator Signature
JARED LEMIEUX

Remarks:

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BMDC 8164.14.00 09/16
01/21/2020 18:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:36
02 Std. Gas	0.082	18:36
03 Room Air	0.000	18:37
04 Std. Gas	0.083	18:37
05 Room Air	0.000	18:37
06 Std. Gas	0.083	18:38
07 Room Air	0.000	18:38

Lot No = 34917080A3
Cyl No = 36
Exp Date = 02/05/2020
County = 08 Oper No. = 130315


Operator Signature
JARED LEMIEUX

Remarks:

*Good to go for now
- will change gas -- 1/22/20*

Form 106-18000

** One closet to door.*