



# ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Demetrius J Jamerson</i>		
Location <i>Newt</i>	Intoxilyzer® Serial Number <i>80-005940</i>	
Gas Lot Number <i>24119080A1</i>	Gas Cylinder Number <i>26</i>	Gas Expiration Date <i>11-05-21</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
A. 0.081 AC  
B. 0.081 AC  
C. 0.080 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i> 5672	Date <i>2-3-20</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>2/5/20</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005940  
Location = NEWT      8164.14.00 09/16  
02/03/2020      15:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:13
02 Std. Gas	0.081	15:13
03 Room Air	0.000	15:14
04 Std. Gas	0.081	15:14
05 Room Air	0.000	15:14
06 Std. Gas	0.080	15:15
07 Room Air	0.000	15:15

Lot No = 24119080A1  
Cyl No = 26  
Exp Date = 11/05/2021  
County = 60      Oper No. = 132358

  
Operator Signature  
DEMETRIUS JAMERSON

Remarks:

gas cylinder install

Form 106-18000