



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Allen, Kathryn M.</i>		
Location <i>Cass County Jail</i>	Intoxilyzer® Serial Number <i>80-006496</i>	
Gas Lot Number <i>17618080A3</i>	Gas Cylinder Number <i>73</i>	Gas Expiration Date <i>8/5/2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. *0.083* AC
B. *0.083* AC
C. *0.083* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Kathryn M. Allen</i>	Date <i>2/20/2020</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Eick</i>	Date <i>2/21/20</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006496
Location = CASS 8164.13.00 06/09
02/20/2020 19:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:50
02 Std. Gas	0.083	19:50
03 Room Air	0.000	19:51
04 Std. Gas	0.083	19:51
05 Room Air	0.000	19:52
06 Std. Gas	0.083	19:52
07 Room Air	0.000	19:53


Lot No = 17618080A3

Cyl No = 73

Exp Date = 08/05/2020

County = 09

Oper No. = 090801



Operator Signature
KATHRYN ALLEN

Remarks:

Form 106-I8000