

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004950 Location: ToxL

A. General Setup and Checks:

- 1.  Diagnostic Tests Pass and Instrument in Ready Mode
- 2.  Breath tube heated
- 3.  Date, time and location code (Level 2,E). Re-set if necessary.
- 4.  Print test (Level 1,P). Sign and attach test record.
- 5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 432 psi Regulator: 450 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

- 1. Wet Calibration Check
  - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.015 AC  
Sim. Ser #: MP3069 Lot #: 201805C Exp. Date: 5.30.20
  - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3069 Lot #: 201902D Exp. Date: 2.20.21

Configure dry gas standard for the following tests (Level 1,S).

- 2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA  
 Display reads "Interferent Detect"
- 3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
- 4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34418080A2 Cyl #: 5 Exp. Date: 2.5.21

Test 1	<u>0.082</u>	Test 4	<u>0.082</u>	Test 7	<u>0.083</u>
Test 2	<u>0.082</u>	Test 5	<u>0.082</u>	Test 8	<u>0.082</u>
Test 3	<u>0.082</u>	Test 6	<u>0.082</u>	Test 9	<u>0.082</u>
Average	<u>0.082</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

---

---

---

Instrument is acceptable to be used in the field.

Charles E. Ee  
Breath Analyst Signature

4/8/20  
Date

NA

NA

Reviewed by \_\_\_\_\_

Date \_\_\_\_\_

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004950  
Location = TOXL                    8164.14.00 09/16  
04/07/2020                            16:48

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            34418080A2  
Standard Cyl #?            5  
Standard Expiration?       02/05/2021  
Oper No?                    666666

Flow Cal. Date:            01/18/2010  
Slope                        774  
Intercept                   -908904

IR Calibration Date:        05/05/2016  
                                  3um                            9um

-----  
0th Coef(\*100):            -13717                        -19495  
1st Coef(\*100):            271916                        140513  
2nd Coef(\*100):            3504                            1503  
H2O adj(mg/l\*10k):        370                            426

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004950  
Location = TOXL      8164.14.00 09/16  
04/07/2020      16:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:55
02 Std. Sol.	0.014	16:56
03 Room Air	0.000	16:56
04 Std. Sol.	0.015	16:57
05 Room Air	0.000	16:58
06 Std. Sol.	0.014	16:58
07 Room Air	0.000	16:59

08 Sim Temp = 34.0°C

Simul Ser No = MP3064  
Std Sol No = 201805C  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.015 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004950  
Location = TOXL      8164.14.00 09/16  
04/08/2020      12:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:36
02 Std. Sol.	0.303	12:37
03 Room Air	0.000	12:37
04 Std. Sol.	0.304	12:38
05 Room Air	0.000	12:38
06 Std. Sol.	0.305	12:39
07 Room Air	0.000	12:40


08 Sim Temp = 34.0°C

Simul Ser No = MP3069

Std Sol No = 201902D

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004950  
Location = TOXL      8164.14.00 09/16  
04/07/2020      16:48

Test	AC	Time
01 Room Air	0.000	16:49
02 *Subject Test	INT*	16:49
03 Room Air	0.000	16:50

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS921456  
Lot No = 34418080A2  
Cyl No = 5  
Expiration Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004950  
Location = TOXL                8164.14.00 09/16  
04/07/2020                        16:50

Test	AC	Time
01 Diagnostic	OK	16:51
02 Room Air	0.000	16:52
03 *Subject Test	RFI*	16:52
04 Room Air	0.000	16:52

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 34418080A2  
Cyl No = 5  
Expiration Date = 02/05/2021  
County = 08                          Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004950  
Location = TOXL      8164.14.00 09/16  
04/08/2020      12:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:42
02 Std. Gas	0.082	12:42
03 Room Air	0.000	12:42
04 Std. Gas	0.082	12:43
05 Room Air	0.000	12:43
06 Std. Gas	0.082	12:44
07 Room Air	0.000	12:44

Lot No = 34418080A2  
Cyl No = 5  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666

  
Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004950  
Location = TOXL      8164.14.00 09/16  
04/08/2020      12:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:45
02 Std. Gas	0.082	12:45
03 Room Air	0.000	12:46
04 Std. Gas	0.082	12:46
05 Room Air	0.000	12:47
06 Std. Gas	0.082	12:47
07 Room Air	0.000	12:47

Lot No = 34418080A2  
Cyl No = 5  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004950  
Location = TOXL      8164.14.00 09/16  
04/08/2020      12:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:48
02 Std. Gas	0.083	12:49
03 Room Air	0.000	12:49
04 Std. Gas	0.082	12:50
05 Room Air	0.000	12:50
06 Std. Gas	0.082	12:50
07 Room Air	0.000	12:51

Lot No = 34418080A2  
Cyl No = 5  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC<sup>CEE</sup> C*

Form 106-I8000