

**INTOXILYZER® 8000 INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-004935

Location: TOXL

**A. General Setup and Checks:**

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 393 psi Regulator: 400 psi

**B. Tests (Sign and attach test records):**

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.015 AC  
Sim. Ser #: MP3064 Lot #: 201805C Exp. Date: 5.30.20
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3069 Lot #: 201902D Exp. Date: 2.20.21

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34418080A2 Cyl #: 5 Exp. Date: 2.5.21  
Test 1 0.078 Test 4 0.079 Test 7 0.079  
Test 2 0.079 Test 5 0.079 Test 8 0.079  
Test 3 0.079 Test 6 0.079 Test 9 0.079  
Average 0.079

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED SIMULATOR RETURN O-RING

Instrument is acceptable to be used in the field.

Charles E. Eh  
Breath Analyst Signature

4.14.20  
Date

NA  
Reviewed by

NA  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004935  
Location = TOXL                    8164.14.00 09/16  
04/14/2020                            15:15

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            34418080A2  
Standard Cyl #?            5  
Standard Expiration?       02/05/2021  
Oper No?                    666666  
  
Flow Cal. Date:            07/12/2012  
Slope                        655  
Intercept                   -510289

IR Calibration Date:       05/20/2015  
                                    3um                            9um

-----  
0th Coef(\*100):        -7871                        -19684  
1st Coef(\*100):        264505                       133108  
2nd Coef(\*100):        2865                         1320  
H2O adj(mg/l\*10k):    200                          298

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004935  
Location = TOXL      8164.14.00 09/16  
04/14/2020      15:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:25
02 Std. Sol.	0.013	15:26
03 Room Air	0.000	15:26
04 Std. Sol.	0.013	15:27
05 Room Air	0.000	15:28
06 Std. Sol.	0.014	15:28
07 Room Air	0.000	15:29

08 Sim Temp = 34.0°C

Simul Ser No = MP3064  
Std Sol No = 201805C  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

LOW AC  
0.015 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004935  
Location = TOXL      8164.14.00 09/16  
04/14/2020      15:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:35
02 Std. Sol.	0.300	15:36
03 Room Air	0.000	15:37
04 Std. Sol.	0.302	15:37
05 Room Air	0.000	15:38
06 Std. Sol.	0.301	15:39
07 Room Air	0.000	15:39

08 Sim Temp = 34.0°C

Simul Ser No = MP3069

Std Sol No = 201902D

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004935  
Location = TOXL      8164.14.00 09/16  
04/14/2020      15:22

Test	AC	Time
01 Room Air	0.000	15:23
02 *Subject Test	INT*	15:23
03 Room Air	0.000	15:24

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERRENT  
Dr. Lic. = ND/DIS921456  
Lot No = 34418080A2  
Cyl No = 5  
Expiration Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10AE + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004935  
Location = TOXL      8164.14.00 09/16  
04/14/2020      15:20

Test	AC	Time
01 Diagnostic	OK	15:21
02 Room Air	0.000	15:21
03 *Subject Test	RFI*	15:22
04 Room Air	0.000	15:22

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 34418080A2  
Cyl No = 5  
Expiration Date = 02/05/2021  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004935  
Location = TOXL      8164.14.00 09/16  
04/14/2020      15:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:41
02 Std. Gas	0.078	15:41
03 Room Air	0.000	15:42
04 Std. Gas	0.079	15:42
05 Room Air	0.000	15:43
06 Std. Gas	0.079	15:43
07 Room Air	0.000	15:44

Lot No = 34418080A2  
Cyl No = 5  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004935  
Location = TOXL      8164.14.00 09/16  
04/14/2020      15:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:44
02 Std. Gas	0.079	15:45
03 Room Air	0.000	15:45
04 Std. Gas	0.079	15:46
05 Room Air	0.000	15:46
06 Std. Gas	0.079	15:47
07 Room Air	0.000	15:47

Lot No = 34418080A2  
Cyl No = 5  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004935  
Location = TOXL      8164.14.00 09/16  
04/14/2020      15:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:48
02 Std. Gas	0.079	15:48
03 Room Air	0.000	15:49
04 Std. Gas	0.079	15:49
05 Room Air	0.000	15:50
06 Std. Gas	0.079	15:50
07 Room Air	0.000	15:51

Lot No = 34418080A2  
Cyl No = 5  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000