

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-003071 Location: Tox L

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 370 psi Regulator: 400 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.015 AC  
Sim. Ser #: MP3064 Lot #: 201805 C Exp. Date: 5.30.20
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3069 Lot #: 201902 D Exp. Date: 2.20.21

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: IC55 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34418080A2 Cyl #: 5 Exp. Date: 2.5.21  
Test 1 0.079 Test 4 0.080 Test 7 0.079  
Test 2 0.080 Test 5 0.079 Test 8 0.079  
Test 3 0.079 Test 6 0.080 Test 9 0.078  
Average 0.079

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Ed  
Breath Analyst Signature

4.17.20  
Date

NA  
Reviewed by

NA  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003071  
Location = TOXL                8164.14.00 09/16  
04/17/2020                      13:12

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\* ()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\* ()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            34418080A2  
Standard Cyl #?            5  
Standard Expiration?       02/05/2021  
Oper No?                    666666

Flow Cal. Date:            07/09/2010  
Slope                      806  
Intercept                  -994737

IR Calibration Date:       04/10/2015  
                                    3um                      9um

-----  
0th Coef(\*100):        -11328                      -10938  
1st Coef(\*100):        268087                      134278  
2nd Coef(\*100):        3093                        1368  
H2O adj(mg/l\*10k):    242                        233

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003071  
Location = TOXL      8164.14.00 09/16  
04/17/2020      13:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:13
02 Std. Sol.	0.014	13:14
03 Room Air	0.000	13:14
04 Std. Sol.	0.014	13:15
05 Room Air	0.000	13:16
06 Std. Sol.	0.014	13:16
07 Room Air	0.000	13:17

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 201805C

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC

0.015 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003071  
Location = TOXL      8164.14.00 09/16  
04/17/2020      13:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:24
02 Std. Sol.	0.294	13:25
03 Room Air	0.000	13:25
04 Std. Sol.	0.295	13:26
05 Room Air	0.000	13:27
06 Std. Sol.	0.295	13:27
07 Room Air	0.000	13:28

08 Sim Temp = 34.0°C

Simul Ser No = MP3069  
Std Sol No = 201902D  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003071  
Location = TOXL      8164.14.00 09/16  
04/17/2020      13:33

Test	AC	Time
01 Room Air	0.000	13:34
02 *Subject Test	INT*	13:34
03 Room Air	0.000	13:35

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS921456  
Lot No = 34418080A2  
Cyl No = 5  
Expiration Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003071  
Location = TOXL                8164.14.00 09/16  
04/17/2020                        13:36

Test	AC	Time
01 Diagnostic	OK	13:37
02 Room Air	0.000	13:38
03 *Subject Test	RFI*	13:38
04 Room Air	0.000	13:38

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 34418080A2  
Cyl No = 5  
Expiration Date = 02/05/2021  
County = 08                          Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003071  
Location = TOXL      8164.14.00 09/16  
04/17/2020      13:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:40
02 Std. Gas	0.079	13:40
03 Room Air	0.000	13:41
04 Std. Gas	0.080	13:41
05 Room Air	0.000	13:42
06 Std. Gas	0.079	13:42
07 Room Air	0.000	13:43

Lot No = 34418080A2  
Cyl No = 5  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003071  
Location = TOXL      8164.14.00 09/16  
04/17/2020      13:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:43
02 Std. Gas	0.080	13:44
03 Room Air	0.000	13:44
04 Std. Gas	0.079	13:45
05 Room Air	0.000	13:45
06 Std. Gas	0.080	13:45
07 Room Air	0.000	13:46

Lot No = 34418080A2  
Cyl No = 5  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003071  
Location = TOXL      8164.14.00 09/16  
04/17/2020      13:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:47
02 Std. Gas	0.079	13:47
03 Room Air	0.000	13:48
04 Std. Gas	0.079	13:48
05 Room Air	0.000	13:49
06 Std. Gas	0.078	13:49
07 Room Air	0.000	13:49

Lot No = 34418080A2  
Cyl No = 5  
Exp Date = 02/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000