

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-002667 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 370 psi Regulator: 375 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.015 AC
Sim. Ser #: MP3064 Lot #: 201805C Exp. Date: 5.30.20
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3069 Lot #: 201902D Exp. Date: 2.20.21

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 05620080A1 Cyl #: 47 Exp. Date: 4.5.22

Test 1 <u>0.082</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.083</u>	Test 8 <u>0.081</u>
Test 3 <u>0.082</u>	Test 6 <u>0.082</u>	Test 9 <u>0.081</u>
Average <u>0.082</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edler
Breath Analyst Signature

4.22.20
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 12:34

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05620080A1
Standard Cyl #? 47
Standard Expiration? 04/05/2022
Oper No? 666666

Flow Cal. Date: 09/10/2007
Slope 693
Intercept -570251

IR Calibration Date: 05/08/2017

	3um	9um
0th Coef(*100):	-16077	-8130
1st Coef(*100):	268518	134907
2nd Coef(*100):	2115	1175
H2O adj(mg/l*10k):	380	220

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 12:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:38
02 Std. Sol.	0.016	12:38
03 Room Air	0.000	12:39
04 Std. Sol.	0.016	12:40
05 Room Air	0.000	12:40
06 Std. Sol.	0.016	12:41
07 Room Air	0.000	12:41

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 201805C

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low Ac
0.015 Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 12:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:43
02 Std. Sol.	0.309	12:44
03 Room Air	0.000	12:44
04 Std. Sol.	0.309	12:45
05 Room Air	0.000	12:45
06 Std. Sol.	0.309	12:46
07 Room Air	0.000	12:47


08 Sim Temp = 34.0°C

Simul Ser No = MP3069

Std Sol No = 201902D

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 12:48

Test	AC	Time
01 Room Air	0.000	12:49
02 *Subject Test	INT*	12:49
03 Room Air	0.000	12:50

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 12:50

Test	AC	Time
01 Diagnostic	OK	12:51
02 Room Air	0.000	12:51
03 *Subject Test	RFI*	12:51
04 Room Air	0.000	12:52

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 12:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:53
02 Std. Gas	0.082	12:53
03 Room Air	0.000	12:54
04 Std. Gas	0.082	12:54
05 Room Air	0.000	12:55
06 Std. Gas	0.082	12:55
07 Room Air	0.000	12:55

Lot No = 05620080A1
Cyl No = 47
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 12:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:57
02 Std. Gas	0.082	12:58
03 Room Air	0.000	12:58
04 Std. Gas	0.083	12:58
05 Room Air	0.000	12:59
06 Std. Gas	0.082	12:59
07 Room Air	0.000	13:00

Lot No = 05620080A1
Cyl No = 47
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 13:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:01
02 Std. Gas	0.082	13:01
03 Room Air	0.000	13:02
04 Std. Gas	0.081	13:02
05 Room Air	0.000	13:02
06 Std. Gas	0.081	13:03
07 Room Air	0.000	13:03

Lot No = 05620080A1
Cyl No = 47
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-18000