

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.0 I: 0.0
B: 45.0 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 979 psi P (gauge): 1000 psi
7. General Setup (Level 3,S,G)
8. Configure Start Test (Level 3,S,T)
9. Network (Level 3,S,N)

E. TESTS (attach all test records):

Configure wet bath simulator standards before running tests (Level 1,S):

1. Zero AC (Level 1,C):
Sim ser #: MP3002 Lot #: Mix of H₂O Exp date: NA
Known value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):
Sim ser #: MP3061 Lot #: 201810D Exp date: 10.24.20
Known value: 0.020 AC
Results: 0.020 AC 0.020 AC 0.021 AC
3. High AC (Level 1,C):
Sim ser #: MP3067 Lot #: 201911B Exp date: 11.5.21
Known value: 0.250 AC
Results: 0.253 AC 0.253 AC 0.254 AC

Configure dry gas standard before running tests (Level 1,S):

4. Ethanol + 0.05% Acetone (Level 1,B):
Sim ser #: DR3451 Lot #: ICS 5 Exp date: NA
Known value: 0.10 AC + 0.05% ACETONE
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):
Sim ser #: DR5188 Lot #: ACE 1 Exp date: NA
Known value: 0.05% ACETONE
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):
Sim ser #: DR3455 Lot #: MET 2 Exp date: NA
Known value: 0.10% METHANOL
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):
Sim ser #: DR5189 Lot #: IPA 2 Exp date: NA
Known value: 0.10% ISOPROPANOL
 Display reads "Interferent Detect"

- 8. Ambient Fail (Level 1,B): Display and test record read correct.
- 9. Invalid Sample (Level 1,B): Display and test record read correct.
- 10. Sequence Aborted (Level 1,B): Display and test record read correct.
- 11. ABA (Level 1,B): Test record reads correct.
- 12. Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13. RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14. Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15. Deficient Samples (CMS mode): Display and test record read correct.
- 16. Improper Sample (CMS mode): Display and test record read correct.
- 17. Difference Too Great (CMS mode): Display and test record read correct.
- 18. CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):

Ave: 0.0816 Std Dev: 0.0005 Rel Std Dev: 0.62%

Maintenance/Repairs/Notes: REPLACED REAL TIME CLOCK BATTERY.
INITIAL INSPECTION DUE TO OPENING UP (TAKING COVER
OFF) TO REPAIR INSTRUMENT. CEE 4.21.20

Instrument is acceptable to be used in the field.


Field Inspector's Signature

4.22.20
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TCXL 8164.14.00 09/16
04/22/2020 10:09

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

32 characters per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

42 characters per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

Current Instrument Setup
Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On MemFull? Yes
of Print Copies? 1
Select Std (D/W/I)? Wet
Standard Value? 0.020
Standard Solution #? 2018100
Sim Serial #? MP3061
Oper No? 666666

Flow Cal. Date: 09/10/2007
Slope: 693
Intercept: -570251

IR Calibration Date: 05/08/2017
3um 9um

0th Coef(*100): -16077 -8130

1st Coef(*100): 268518 134907

2nd Coef(*100): 2115 1175

H2O adj.(ng/l*10K): 380 220

***** Test End *****



Operator Signature
CHARLES EDER

Remarks: INTERNAL

PRINT TEST

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 10:13

Test	AC	Time
01 Room Air	0.000	10:13
02 *Subject Test	REF*	10:14
03 Room Air	0.000	10:14

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = R KEY REFUSAL
Dr. Lic. = ND/DIS921456
Simul Ser No = MP3061
Std Sol No = 201810D
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: "R" KEY REFUSAL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 10:15

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

DIAGNOSTIC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 10:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Sol.	0.000	10:24
03 Room Air	0.000	10:24
04 Std. Sol.	0.000	10:25
05 Room Air	0.000	10:26
06 Std. Sol.	0.000	10:26
07 Room Air	0.000	10:27

08 Sim Temp = 34.0°C

Simul Ser No = MP3002
Std Sol No = MILLIQ H2O
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ZERO AC
0.000 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 10:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:30
02 Std. Sol.	0.020	10:31
03 Room Air	0.000	10:31
04 Std. Sol.	0.020	10:32
05 Room Air	0.000	10:32
06 Std. Sol.	0.021	10:33
07 Room Air	0.000	10:33

08 Sim Temp = 34.0°C

Simul Ser No = MP2061
Std Sol No = 201810D
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LOW AC*
 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 10:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:43
02 Std. Sol.	0.253	10:44
03 Room Air	0.000	10:45
04 Std. Sol.	0.253	10:45
05 Room Air	0.000	10:46
06 Std. Sol.	0.254	10:47
07 Room Air	0.000	10:47

08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 201911B

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.250 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 10:52

Test	AC	Time
01 Room Air	0.000	10:52
02 *Subject Test	INT*	10:53
03 Room Air	0.000	10:53

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 10:54

Test	AC	Time
01 Room Air	0.000	10:55
02 *Subject Test	INT*	10:55
03 Room Air	0.000	10:56

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 10:56

Test	AC	Time
01 Room Air	0.000	10:57
02 *Subject Test	INT*	10:57
03 Room Air	0.000	10:58

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10% METHANOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 10:59

Test	AC	Time
01 Room Air	0.000	11:00
02 *Subject Test	INT*	11:00
03 Room Air	0.000	11:01

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10% ISOPROPANOL
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 11:02

Test	AC	Time
01 Room Air	AMB*	11:03
02 Room Air	0.000	11:03

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = AMBIENT FAIL
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

AMBIENT FAIL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 11:06

Test	AC	Time
01 Room Air	0.000	11:07
02 Invalid Sample	X.XXX	11:07
03 Room Air	0.000	11:08

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INVALID SAMPLE
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INVALID SAMPLE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 11:08

Test	AC	Time
01 Room Air	0.000	11:09
02 *Subject Test	ABT*	11:09
03 Room Air	0.000	11:09

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = SEQ ABORT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: SEQUENCE ABORTED

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 11:11

Test	AC	Time
01 Room Air	0.000	11:12
02 Subject Test 1	0.000	11:12
03 Room Air	0.000	11:13
04 Reported AC	0.000	11:12

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = ABA MODE
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ABA MODE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 11:13

Test	AC	Time
01 Diagnostic	OK	11:14
02 Room Air	0.000	11:14
03 Subject Test 1	0.053	11:15
04 Room Air	0.000	11:17
05 Std. Gas	0.088*	11:18
06 Room Air	0.000	11:19

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = CAL CHK OUT TOL
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
OUT OF TOLERANCE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 11:20

Test	AC	Time
01 Diagnostic	OK	11:21
02 Room Air	0.000	11:21
03 *Subject Test	RFI*	11:21
04 Room Air	0.000	11:22

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 11:22

Test	AC	Time
01 Diagnostic	OK	11:23
02 Room Air	0.000	11:23
03 *Subject Test	0.000*	11:26
04 Room Air	0.000	11:29
05 Std. Gas	0.081	11:30
06 Room Air	0.000	11:31
07 *Subject Test	0.000*	11:34
08 Room Air	0.000	11:35


No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = DEFICIENT SAMP
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

DEFICIENT SAMPLES

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 11:35

Test	AC	Time
01 Diagnostic	OK	11:36
02 Room Air	0.000	11:37
03 Reference	IPS*	11:37
04 Room Air	0.000	11:37

*Improper Sample

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = IMPROPER SAMPLE
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

IMPROPER SAMPLE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 11:38

Test	AC	Time
01 Diagnostic	OK	11:39
02 Room Air	0.000	11:39
03 Subject Test 1	0.000*	11:39
04 Room Air	0.000	11:42
05 Std. Gas	0.081	11:43
06 Room Air	0.000	11:44
07 Subject Test 2	0.053*	11:44
08 Room Air	0.000	11:45

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = DIFF TOO GREAT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

DIFFERENCE

TOO GREAT

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 11:46

Test	AC	Time
01 Diagnostic	OK	11:46
02 Room Air	0.000	11:47
03 Subject Test 1	0.052	11:47
04 Room Air	0.000	11:50
05 Std. Gas	0.081	11:51
06 Room Air	0.000	11:52
07 Subject Test 2	0.052	11:52
08 Room Air	0.000	11:53
09 Reported AC	0.052	11:52

Difference OK
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = CMS MODE
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

CMS MODE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.14.00 09/16
04/22/2020 11:54

Test	AC	Time
01 Room Air	0.000	11:54
02 Std. Gas	0.081	11:55
03 Room Air	0.000	11:55
04 Std. Gas	0.081	11:55
05 Room Air	0.000	11:56
06 Std. Gas	0.081	11:56
07 Room Air	0.000	11:57
08 Std. Gas	0.082	11:57
09 Room Air	0.000	11:58
10 Std. Gas	0.081	11:58
11 Room Air	0.000	11:59
12 Std. Gas	0.082	11:59
13 Room Air	0.000	11:59
14 Std. Gas	0.081	12:00
15 Room Air	0.000	12:00
16 Std. Gas	0.082	12:01
17 Room Air	0.000	12:01
18 Std. Gas	0.081	12:01
19 Room Air	0.000	12:02
20 Std. Gas	0.082	12:02
21 Room Air	0.000	12:03
22 Std. Gas	0.081	12:03
23 Room Air	0.000	12:04
24 Std. Gas	0.082	12:04
25 Room Air	0.000	12:04
26 Std. Gas	0.082	12:05
27 Room Air	0.000	12:05
28 Std. Gas	0.082	12:06
29 Room Air	0.000	12:06
30 Std. Gas	0.082	12:06
31 Room Air	0.000	12:07
32 Std. Gas	0.082	12:07
33 Room Air	0.000	12:08
34 Std. Gas	0.082	12:08
35 Room Air	0.000	12:09
36 Std. Gas	0.082	12:09
37 Room Air	0.000	12:09
38 Std. Gas	0.082	12:10
39 Room Air	0.000	12:10
40 Std. Gas	0.081	12:11
41 Room Air	0.000	12:11
42 Cal Check Stats:		
43 Average	0.0816	
44 Std Dev	0.0005	
45 Rel Std Dev	0.62%	



Operator Signature
CHARLES EDER

Remarks: STABILITY TEST

Form 106-I8000

LOT: 24119080A1

CYL: 9

EXP: 11.5.21