

INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-003066

A. GENERAL Setup and Checks:

80-003066

- Power On
 5 min. countdown during initial warm-up
 1 min. countdown to begin test after standby mode
- Display Intensity (Ctrl + ↑ or ↓)
- Password Security:
 Level 1 (Operator) Level 2 (Field Inspector) Level 3 (Lab Personnel)
- Printer Setup (Level 3,S,P)
- Forms Load (CMI, Inc., CoBrA Software)

B. LEVEL 1 Setup and Checks:

- Internal Printer Test and Paper Feed (Level 1,P). Attach record.
- External Printer Test (Level 1,P). Attach record.
- Barcode Reader (Level 1,B):
 Operator ID ND License R key: Refusal. Attach record.

C. LEVEL 2 Setup and Checks:

- Continuous Air Blank (Level 2,A)
- Diagnostic (Level 2,D). Attach record.
- Prelim Data Entry (Level 2,E)
- Ser Number/Version Info (Level 2,V):
Inst Ser#: 80-003066 DSP Ver: 8164.14
Core Ver: 1362.14 CPLD Ver: 7

D. LEVEL 3 Setup and Checks:

- DVM Monitor (Level 3,D,D): 3: 12876 s: 1 c: 47.00
9: 13533 s: 1 B: 45 0
(RFI indicator with keyed radio): 1
- Function (F) Keys: : F5 – Tone F9 - Fan
 F6 – Green LED F10 – Pump
 F7 – Red LED F11 – Solenoid
- Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 957 CF: 1.05
- Flow Monitor (Level 3,D,F): Flow Sensor, R: 140 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 46.5 I: 0.0
B: 45.0 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 350 psi P (gauge): 375 psi
7. General Setup (Level 3,S,G)
8. Configure Start Test (Level 3,S,T)
9. Network (Level 3,S,N)

E. TESTS (attach all test records):

Configure wet bath simulator standards before running tests (Level 1,S):

1. Zero AC (Level 1,C): 02 CEE
Sim ser #: MP3067 Lot #: Milli H₂O Exp date: NA
Known value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):
Sim ser #: MP3061 Lot #: 2018100 Exp date: 10.24.20
Known value: 0.020 AC
Results: 0.021 AC 0.020 AC 0.020 AC
3. High AC (Level 1,C):
Sim ser #: MP3067 Lot #: 201911B Exp date: 11.5.21
Known value: 0.250 AC
Results: 0.249 AC 0.249 AC 0.249 AC

Configure dry gas standard before running tests (Level 1,S):

4. Ethanol + 0.05% Acetone (Level 1,B):
Sim ser #: DR3451 Lot #: ICS 5 Exp date: NA
Known value: 0.10 AC + 0.05% ACETONE
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):
Sim ser #: DR5188 Lot #: ACE 2 Exp date: NA
Known value: 0.05% ACETONE
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):
Sim ser #: DR3455 Lot #: MET 2 Exp date: NA
Known value: 0.10% METHANOL
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):
Sim ser #: DR5189 Lot #: IPA 2 Exp date: NA
Known value: 0.10% ISOPROPANOL
 Display reads "Interferent Detect"

- 8. Ambient Fail (Level 1,B): Display and test record read correct.
- 9. Invalid Sample (Level 1,B): Display and test record read correct.
- 10. Sequence Aborted (Level 1,B): Display and test record read correct.
- 11. ABA (Level 1,B): Test record reads correct.
- 12. Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13. RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14. Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15. Deficient Samples (CMS mode): Display and test record read correct.
- 16. Improper Sample (CMS mode): Display and test record read correct.
- 17. Difference Too Great (CMS mode): Display and test record read correct.
- 18. CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):

Ave: 0.0800 Std Dev: 0.0007 Rel Std Dev: 0.86%

Maintenance/Repairs/Notes: REPLACED REAL TIME CLOCK BATTERY.
INITIAL INSPECTION DUE TO OPENING UP INSTRUMENT AND
PERFORMING REPAIR.

Instrument is acceptable to be used in the field.


Field Inspector's Signature

4.29.20
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 11:45

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMNPO

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNOPQRSTU

24 characters per line
abcdefghijklmnopqrstuvw
ABCDEFGHIJKLMNOPQRSTUWVX

32 characters per line
abcdefghijklmnopqrstuvwxy123456
ABCDEFGHIJKLMNOPQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuvwxy1234567890abcdef
ABCDEFGHIJKLMNOPQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup
Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Menfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 5
Standard Expiration? 02/05/2021
Oper No? 666666

Flow Cal. Date: 07/29/2010
Slope 806
Intercept -895033

IR Calibration Date: 07/29/2010
Sum Sum

0th Coef(*100):	-14075	-5808
1st Coef(*100):	272113	138204
2nd Coef(*100):	1637	986
H2O adj(mg/l*10K):	289	148

***** Test End *****



Operator Signature
CHARLES EDER

Remarks: INTERNAL
PRINTER
TEST
Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 13:17

Test	AC	Time
01 Room Air	0.000	13:18
02 *Subject Test	REF*	13:18
03 Room Air	0.000	13:19

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = R KEY REFUSAL
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 5
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: "R" KEY REFUSAL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 13:20

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

DIAGNOSTIC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 13:40

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:40
02 Std. Sol.	0.000	13:41
03 Room Air	0.000	13:42
04 Std. Sol.	0.000	13:42
05 Room Air	0.000	13:43
06 Std. Sol.	0.000	13:44
07 Room Air	0.000	13:44

08 Sim Temp = 34.0°C

Simul Ser No = MP3002
Std Sol No = MILLIQ H2O
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ZERO AC
0.000 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 13:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:47
02 Std. Sol.	0.021	13:48
03 Room Air	0.000	13:49
04 Std. Sol.	0.020	13:49
05 Room Air	0.000	13:50
06 Std. Sol.	0.020	13:51
07 Room Air	0.000	13:51

08 Sim Temp = 34.0°C

Simul Ser No = MP3061
Std Sol No = 201810D
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 13:54

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:54
02 Std. Sol.	0.249	13:55
03 Room Air	0.000	13:56
04 Std. Sol.	0.249	13:56
05 Room Air	0.000	13:57
06 Std. Sol.	0.249	13:58
07 Room Air	0.000	13:58

08 Sim Temp = 34.0°C

Simul Ser No = MP3067
Std Sol No = 201911B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.250 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 14:02

Test	AC	Time
01 Room Air	0.000	14:02
02 *Subject Test	INT*	14:03
03 Room Air	0.000	14:03

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 14:05

Test	AC	Time
01 Room Air	0.000	14:06
02 *Subject Test	INT*	14:06
03 Room Air	0.000	14:07

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 14:08

Test	AC	Time
01 Room Air	0.000	14:09
02 *Subject Test	INT*	14:10
03 Room Air	0.000	14:10

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check

0.10% METHANOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 14:12

Test	AC	Time
01 Room Air	0.000	14:12
02 *Subject Test	INT*	14:13
03 Room Air	0.000	14:14

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10% ISOPROPANOL

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 14:17

Test	AC	Time
01 Room Air	AMB*	14:18
02 Room Air	0.000	14:18

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = AMBIENT FAIL
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

AMBIENT FAIL

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 14:19

Test	AC	Time
01 Room Air	0.000	14:19
02 Invalid Sample	X.XXX	14:20
03 Room Air	0.000	14:20

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INVALID SAMPLE
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INVALID SAMPLE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 14:21

Test	AC	Time
01 Room Air	0.000	14:22
02 *Subject Test	ABT*	14:22
03 Room Air	0.000	14:23

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = SEQ ABORTED
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: SEQUENCE ABORTED

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 14:32

Test	AC	Time
01 Room Air	0.000	14:33
02 Subject Test 1	0.052	14:34
03 Room Air	0.000	14:34
04 Reported AC	0.052	14:34

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = ABA MODE
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ABA MODE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 14:36

Test	AC	Time
01 Diagnostic	OK	14:36
02 Room Air	0.000	14:37
03 Subject Test 1	0.052	14:37
04 Room Air	0.000	14:40
05 Std. Gas	0.087*	14:41
06 Room Air	0.000	14:41

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = CAL CHK OUT TOL
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
OUT OF TOLERANCE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 14:43

Test	AC	Time
01 Diagnostic	OK	14:44
02 Room Air	0.000	14:44
03 *Subject Test	RFI*	14:44
04 Room Air	0.000	14:45

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 14:46

Test	AC	Time
01 Diagnostic	OK	14:46
02 Room Air	0.000	14:47
03 *Subject Test	0.000*	14:50
04 Room Air	0.000	14:52
05 Std. Gas	0.080	14:53
06 Room Air	0.000	14:55
07 *Subject Test	0.000*	14:58
08 Room Air	0.000	14:58

No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = DEFICIENT SAMP
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

DEFICIENT
SAMPLES

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 15:03

Test	AC	Time
01 Diagnostic	OK	15:04
02 Room Air	0.000	15:04
03 Reference	IPS*	15:04
04 Room Air	0.000	15:05

*Improper Sample

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = IMPROPER SAMPLE
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *IMPROPER SAMPLE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 15:06

Test	AC	Time
01 Diagnostic	OK	15:07
02 Room Air	0.000	15:07
03 Subject Test 1	0.052*	15:08
04 Room Air	0.000	15:10
05 Std. Gas	0.080	15:11
06 Room Air	0.000	15:12
07 Subject Test 2	0.104*	15:13
08 Room Air	0.000	15:14

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = DIFF TOO GREAT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *DIFFERENCE*
 TOO GREAT

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 15:16


Test	AC	Time
01 Diagnostic	OK	15:17
02 Room Air	0.000	15:17
03 Subject Test 1	0.102	15:18
04 Room Air	0.000	15:20
05 Std. Gas	0.079	15:21
06 Room Air	0.000	15:22
07 Subject Test 2	0.102	15:23
08 Room Air	0.000	15:24
09 Reported AC	0.102	15:23

Difference OK
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = CMS MODE
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

CMS MODE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = TOXL 8164.14.00 09/16
04/29/2020 15:25

Test	AC	Time
01 Room Air	0.000	15:26
02 Std. Gas	0.079	15:26
03 Room Air	0.000	15:27
04 Std. Gas	0.080	15:27
05 Room Air	0.000	15:28
06 Std. Gas	0.080	15:28
07 Room Air	0.000	15:29
08 Std. Gas	0.080	15:29
09 Room Air	0.000	15:30
10 Std. Gas	0.079	15:30
11 Room Air	0.000	15:30
12 Std. Gas	0.080	15:31
13 Room Air	0.000	15:31
14 Std. Gas	0.080	15:32
15 Room Air	0.000	15:32
16 Std. Gas	0.080	15:33
17 Room Air	0.000	15:33
18 Std. Gas	0.080	15:33
19 Room Air	0.000	15:34
20 Std. Gas	0.079	15:34
21 Room Air	0.000	15:35
22 Std. Gas	0.081	15:35
23 Room Air	0.000	15:36
24 Std. Gas	0.080	15:36
25 Room Air	0.000	15:37
26 Std. Gas	0.081	15:37
27 Room Air	0.000	15:38
28 Std. Gas	0.081	15:38
29 Room Air	0.000	15:38
30 Std. Gas	0.080	15:39
31 Room Air	0.000	15:39
32 Std. Gas	0.080	15:40
33 Room Air	0.000	15:40
34 Std. Gas	0.080	15:41
35 Room Air	0.000	15:41
36 Std. Gas	0.079	15:41
37 Room Air	0.000	15:42
38 Std. Gas	0.079	15:42
39 Room Air	0.000	15:43
40 Std. Gas	0.081	15:43
41 Room Air	0.000	15:44
42 Cal Check Stats:		
43 Average	0.0800	
44 Std Dev	0.0007	
45 Rel Std Dev	0.86%	



Operator Signature
CHARLES EDER

Remarks:

STABILITY TEST

Form 106-I8000

LOT: 05620080 A1

CYL: 47

EXP: 4/5/22