

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005941

Location: ToxL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 353 psi Regulator: 375 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.015 AC  
Sim. Ser #: MP3064 Lot #: 201815C Exp. Date: 5-30-20
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3069 Lot #: 201902D Exp. Date: 2-20-21

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: IC55 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 05620080 A1 Cyl #: 47 Exp. Date: 4-5-22

Test 1	<u>0.080</u>	Test 4	<u>0.082</u>	Test 7	<u>0.082</u>
Test 2	<u>0.080</u>	Test 5	<u>0.081</u>	Test 8	<u>0.082</u> CEE 1
Test 3	<u>0.080</u>	Test 6	<u>0.081</u>	Test 9	<u>0.082</u>
Average	<u>0.081</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

CLASSROOM ONLY UNTIL EVALUATED/REPAIRED. AUTO CAL LOW.

Instrument is acceptable to be used in the field.

Charles E. Ed  
Breath Analyst Signature

5/4/20  
Date

NA  
Reviewed by

NA  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005941  
Location = TOXL                      8164.14.00 09/16  
05/04/2020                              14:27

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%\$^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%\$^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                      Enabled  
Start Test Sequence:                    DABACABA  
Display Prelim Rslt?                    Yes  
Display Third Digit?                    Yes  
Inhib Printer(Y/N)?                    No  
Display Volume?                        No  
Disable On Memfull?                    Yes  
# of Print Copies?                      1  
Select Std (D/W/I)?                    Dry  
Standard Value?                        0.080  
Standard Lot #?                         05620080A1  
Standard Cyl #?                         47  
Standard Expiration?                    04/05/2022  
Oper No?                                666666

Flow Cal. Date:                        05/30/2018  
Slope                                    618  
Intercept                                -570330

IR Calibration Date:                    05/30/2018  
    3um                                      9um

-----  
0th Coef (\*100):                        -13785                                      -15807  
1st Coef (\*100):                        281869                                      139821  
2nd Coef (\*100):                        2733                                         1395  
H2O adj (mg/l\*10k):                      308                                         362

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005941  
Location = TOXL      8164.14.00 09/16  
05/04/2020      14:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:34
02 Std. Sol.	0.014	14:36
03 Room Air	0.000	14:36
04 Std. Sol.	0.015	14:37
05 Room Air	0.000	14:37
06 Std. Sol.	0.014	14:38
07 Room Air	0.000	14:39

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 201805C

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC  
0.015 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005941  
Location = TOXL      8164.14.00 09/16  
05/04/2020      14:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:40
02 Std. Sol.	0.302	14:41
03 Room Air	0.000	14:41
04 Std. Sol.	0.305	14:42
05 Room Air	0.000	14:43
06 Std. Sol.	0.304	14:43
07 Room Air	0.000	14:44

08 Sim Temp = 34.0°C

Simul Ser No = MP3069  
Std Sol No = 201902D  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*HIGH AC*  
*0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005941  
Location = TOXL                    8164.14.00 09/16  
05/04/2020                            14:46

Test	AC	Time
01 Room Air	0.000	14:47
02 *Subject Test	INT*	14:47
03 Room Air	0.000	14:48

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS921456  
Lot No = 05620080A1  
Cyl No = 47  
Expiration Date = 04/05/2022  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005941  
Location = TOXL                    8164.14.00 09/16  
05/04/2020                            14:49

Test	AC	Time
01 Diagnostic	OK	14:50
02 Room Air	0.000	14:50
03 *Subject Test	RFI*	14:51
04 Room Air	0.000	14:51

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 05620080A1  
Cyl No = 47  
Expiration Date = 04/05/2022  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005941  
Location = TOXL      8164.14.00 09/16  
05/04/2020      14:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:52
02 Std. Gas	0.080	14:53
03 Room Air	0.000	14:53
04 Std. Gas	0.080	14:54
05 Room Air	0.000	14:54
06 Std. Gas	0.080	14:55
07 Room Air	0.000	14:55

Lot No = 05620080A1  
Cyl No = 47  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005941  
Location = TOXL      8164.14.00 09/16  
05/04/2020      14:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:56
02 Std. Gas	0.082	14:56
03 Room Air	0.000	14:57
04 Std. Gas	0.081	14:57
05 Room Air	0.000	14:58
06 Std. Gas	0.081	14:58
07 Room Air	0.000	14:59

Lot No = 05620080A1  
Cyl No = 47  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005941  
Location = TOXL      8164.14.00 09/16  
05/04/2020      14:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:00
02 Std. Gas	0.082	15:00
03 Room Air	0.000	15:00
04 Std. Gas	0.081	15:01
05 Room Air	0.000	15:01
06 Std. Gas	0.082	15:02
07 Room Air	0.000	15:02

Lot No = 05620080A1  
Cyl No = 47  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000