



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

|  |   |
|--|---|
| Serial Number<br><b>80-005360</b>  | Instrument Location<br><b>Grand Forks Police Department</b> |
| Reason for Install/Repair  |   |
| <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change<br><input type="checkbox"/> Other (Specify) _____ |   |

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): Yearly check by state
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

|   |                         |
|---|-------------------------|
| Field Inspector Signature<br><i>Jon Lantz</i> | Date<br><b>05-08-20</b> |
|---|-------------------------|

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

|   |                                    |
|---|------------------------------------|
| Reviewed/Certified By<br><i>Charles E. Ed</i> | Certified Date<br><b>5.13.2020</b> |
|---|------------------------------------|

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005360  
Location = GFPD                    8164.14.00 09/16  
05/08/2020                            16:00

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            34418080A2  
Standard Cyl #?            59  
Standard Expiration?       02/05/2021  
Oper No?                    039947

Flow Cal. Date:            05/07/2020  
Slope                        645  
Intercept                   -464870

IR Calibration Date:       05/07/2020

|                    | 3um    | 9um    |
|--------------------|--------|--------|
| 0th Coef(*100):    | -16059 | -24146 |
| 1st Coef(*100):    | 273073 | 148552 |
| 2nd Coef(*100):    | 593    | 704    |
| H2O adj(mg/l*10k): | 145    | 249    |

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
JONATHAN LAMPI

Remarks:

Form 106-I8000

ACA TEST


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005360  
Location = GFFD      8164.14.00 09/16  
05/08/2020      16:01

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 16:02 |
| 02 Std. Gas | 0.079 | 16:02 |
| 03 Room Air | 0.000 | 16:03 |
| 04 Std. Gas | 0.081 | 16:03 |
| 05 Room Air | 0.000 | 16:04 |
| 06 Std. Gas | 0.079 | 16:04 |
| 07 Room Air | 0.000 | 16:05 |

Lot No = 34418080A2  
Cyl No = 59  
Exp Date = 02/05/2021  
County = 18      Oper No. = 039947

  
Operator Signature  
JONATHAN LAMPI

Remarks:

Form 106-I8000

Print test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005360  
Location = GFPD      8164.14.00 09/16  
05/08/2020      16:06

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | RFI*  | 16:06 |
| 02 Room Air | 0.000 | 16:07 |

\*Invalid Test  
Inhibited - RFI

Lot No = 34418080A2  
Cyl No = 59  
Exp Date = 02/05/2021  
County = 18      Oper No. = 039947

  
Operator Signature  
JONATHAN LAMPI

Remarks:

RFI test      Form 106-I8000