



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-004184	Instrument Location Divide Co. SO Intake/Processing
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

*CEE
5.26.20*

Field Inspector Signature 	Date 5/26/20
-------------------------------	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date 5/26/2020
---------------------------	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = CROS 8164.14.00 09/16
05/21/2020 10:52

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 32
Standard Expiration? 11/05/2021
Oper No? 130230


Flow Cal. Date: 07/30/2010
Slope 804
Intercept -925008

IR Calibration Date: 12/07/2015

	3um	9um
0th Coef (*100):	-15679	-18680
1st Coef (*100):	263558	136948
2nd Coef (*100):	2594	1146
H2O adj (mg/l*10k):	405	327

***** Printer Test End *****

The undersigned, having custody of the original record, certifies that the information hereon is a true and correct copy of the original document maintained as part of the activities of this office.

Signature:  Date: 5/21/20


Operator Signature
JEREMY GROHS

Remarks: *Good Test*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = CROS 8164.14.00 09/16
05/21/2020 11:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:13
02 Std. Gas	0.078	11:14
03 Room Air	0.000	11:14
04 Std. Gas	0.078	11:15
05 Room Air	0.000	11:15
06 Std. Gas	0.078	11:16
07 Room Air	0.000	11:16


Lot No = 24119080A1
Cyl No = 32
Exp Date = 11/05/2021
County = 12 Oper No. = 130230


Operator Signature
JEREMY GROHS

Remarks: *Good Test*

Form 106-I8000

The undersigned, having custody of the original record, certifies that the information herein is a true and correct copy of the original document maintained as part of the activities of this office.


Signature
Date *5/21/20*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = CROS 8164.14.00 09/16
05/21/2020 11:18

Test	AC	Time
01 Room Air	0.000	11:20
02 *Subject Test	RFI*	11:20
03 Room Air	0.000	11:21

*Invalid Test
Inhibited - RFI


Sub Name = TEST, TEST TEST
Sub DOB = 01/01/1999
Sub Sex = Male Weight = 200
Test = OTH Cit = 9999999
Dr. Lic. = ND/GRO805866
Lot No = 24119080A1
Cyl No = 32
Expiration Date = 11/05/2021
County = 12 Oper No. = 130230


Operator Signature
JEREMY GROHS

Remarks: *Good Test*

Form 106-I8000

The undersigned, having custody of the original record, certifies that the information herein is a true and correct copy of the original document maintained as part of the activities of this office.


Signature
Date *5/21/2020*