



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
SFN 59281 (06/2018)

Serial Number <b>80-005958</b>	Instrument Location <b>BE PD</b>
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): None
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i> 1745	Date <b>5-22-20</b>
------------------------------------------------------	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>[Signature]</i>	Certified Date <b>5/28/2020</b>
---------------------------------------------	------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005958  
Location = BEPD                    8164.14.00 09/16  
05/22/2020                            16:53

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            24119080A1  
Standard Cyl #?            20  
Standard Expiration?       11/05/2021  
Oper No?                    131958

Flow Cal. Date:            06/21/2013  
Slope                        716  
Intercept                   -760436

IR Calibration Date:       04/07/2015  
                                  3um                            9um

-----  
0th Coef(\*100):           -17179                       -17694  
1st Coef(\*100):           259328                       130890  
2nd Coef(\*100):           3473                           1654  
H2O adj(mg/l\*10k):       425                            422

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
ADAM BAKER

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005958  
Location = BEPD                    8164.14.00 09/16  
05/22/2020                            16:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:57
02 Std. Gas	0.079	16:57
03 Room Air	0.000	16:58
04 Std. Gas	0.079	16:58
05 Room Air	0.000	16:59
06 Std. Gas	0.080	16:59
07 Room Air	0.000	16:59

Lot No = 24119080A1  
Cyl No = 20  
Exp Date = 11/05/2021  
County = 29                            Oper No. = 131958

  
\_\_\_\_\_  
Operator Signature  
ADAM BAKER

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005958  
Location = BEPD                8164.14.00 09/16  
05/22/2020                      17:02

Test	AC	Time
01 Room Air	0.000	17:04
02 *Subject Test	RFI*	17:04
03 Room Air	0.000	17:05

\*Invalid Test  
Inhibited - RFI

Sub Name = BAKER, ADAM JAMES  
Sub DOB = 06/29/1982  
Sub Sex = Male                    Weight = 190  
Test = DUI                        Cit = TEST  
Dr. Lic. = ND/BAK823805  
Lot No = 24119080A1  
Cyl No = 20  
Expiration Date = 11/05/2021  
County = 29                        Oper No. = 131958



Operator Signature  
ADAM BAKER

Remarks:

Form 106-I8000