

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006491

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 351 psi Regulator: 375 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3064 Lot #: 20070 Exp. Date: 2.13.22
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3069 Lot #: 201902D Exp. Date: 2.20.21

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 05L20080A1 Cyl #: 47 Exp. Date: 4.5.22
Test 1 0.081 Test 4 0.082 Test 7 0.082
Test 2 0.082 Test 5 0.082 Test 8 0.082
Test 3 0.082 Test 6 0.082 Test 9 0.081
Average 0.082

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edin
Breath Analyst Signature

5.27.20
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
05/27/2020 15:36

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05620080A1
Standard Cyl #? 47
Standard Expiration? 04/05/2022
Oper No? 666666

Flow Cal. Date: 05/22/2018
Slope 639
Intercept -618311

IR Calibration Date: 05/23/2018
 3um 9um

0th Coef(*100): -23641 -17066
1st Coef(*100): 264855 137703
2nd Coef(*100): 2663 1471
H2O adj(mg/l*10k): 556 451

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
05/27/2020 15:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:38
02 Std. Sol.	0.017	15:39
03 Room Air	0.000	15:40
04 Std. Sol.	0.018	15:40
05 Room Air	0.000	15:41
06 Std. Sol.	0.018	15:42
07 Room Air	0.000	15:42

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 20070

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
05/27/2020 15:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:49
02 Std. Sol.	0.298	15:50
03 Room Air	0.000	15:51
04 Std. Sol.	0.299	15:52
05 Room Air	0.000	15:53
06 Std. Sol.	0.298	15:53
07 Room Air	0.000	15:54

08 Sim Temp = 34.0°C

Simul Ser No = MP3069

Std Sol No = 201902D

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
05/27/2020 15:58

Test	AC	Time
01 Room Air	0.000	15:58
02 *Subject Test	INT*	15:59
03 Room Air	0.000	16:00

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
05/27/2020 16:00

Test	AC	Time
01 Diagnostic	OK	16:01
02 Room Air	0.000	16:01
03 *Subject Test	RFI*	16:01
04 Room Air	0.000	16:02

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
05/27/2020 16:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:02
02 Std. Gas	0.081	16:03
03 Room Air	0.000	16:03
04 Std. Gas	0.082	16:04
05 Room Air	0.000	16:04
06 Std. Gas	0.082	16:05
07 Room Air	0.000	16:05

Lot No = 05620080A1
Cyl No = 47
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
05/27/2020 16:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:07
02 Std. Gas	0.082	16:07
03 Room Air	0.000	16:08
04 Std. Gas	0.082	16:08
05 Room Air	0.000	16:09
06 Std. Gas	0.082	16:09
07 Room Air	0.000	16:10

Lot No = 05620080A1
Cyl No = 47
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
05/27/2020 16:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:11
02 Std. Gas	0.082	16:12
03 Room Air	0.000	16:12
04 Std. Gas	0.082	16:13
05 Room Air	0.000	16:13
06 Std. Gas	0.081	16:14
07 Room Air	0.000	16:14

Lot No = 05620080A1
Cyl No = 47
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000