

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005953

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 1000 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP3061 Lot #: 201810D Exp. Date: 10.24.20
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3069 Lot #: 201902D Exp. Date: 2.20.21

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: FCS6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 05620 08041 Cyl #: 34 Exp. Date: 4.5.22

Test 1 <u>0.081</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.082</u>	Test 8 <u>0.082</u>
Test 3 <u>0.081</u>	Test 6 <u>0.082</u>	Test 9 <u>0.081</u>
Average <u>0.082</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION  
TANK SENSOR CALIBRATION

Instrument is acceptable to be used in the field.

Charles E. Ede  
Breath Analyst Signature

6.4.20  
Date

NA  
Reviewed by

NA  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005953  
Location = TOXL                    8164.14.00 09/16  
06/04/2020                            12:44

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            05620080A1  
Standard Cyl #?            34  
Standard Expiration?       04/05/2022  
Oper No?                    666666

Flow Cal. Date:            06/16/2015  
Slope                      678  
Intercept                  -668987

IR Calibration Date:       04/30/2015  
                                  3um                    9um

-----  
0th Coef(\*100):            -13435                -19950  
1st Coef(\*100):            257165                129445  
2nd Coef(\*100):            3995                   1631  
H2O adj(mg/l\*10k):        505                    548

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.14.00 09/16  
06/04/2020      12:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:48
02 Std. Sol.	0.018	12:49
03 Room Air	0.000	12:49
04 Std. Sol.	0.019	12:50
05 Room Air	0.000	12:50
06 Std. Sol.	0.018	12:51
07 Room Air	0.000	12:51

08 Sim Temp = 34.0°C

Simul Ser No = MP3061  
Std Sol No = 201810D  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Low AC*  
*0.020 AC*

Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.14.00 09/16  
06/04/2020      12:58

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:58
02 Std. Sol.	0.298	12:59
03 Room Air	0.000	12:59
04 Std. Sol.	0.298	13:00
05 Room Air	0.000	13:01
06 Std. Sol.	0.298	13:01
07 Room Air	0.000	13:02


08 Sim Temp = 34.0°C

Simul Ser No = MP3069

Std Sol No = 201902D

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC

0.300 AC

Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.14.00 09/16  
06/04/2020      13:04

Test	AC	Time
01 Room Air	0.000	13:05
02 *Subject Test	INT*	13:05
03 Room Air	0.000	13:06

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS921456  
Lot No = 05620080A1  
Cyl No = 34  
Expiration Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.14.00 09/16  
06/04/2020      13:06

Test	AC	Time
01 Diagnostic	OK	13:07
02 Room Air	0.000	13:07
03 *Subject Test	RFI*	13:07
04 Room Air	0.000	13:08

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 05620080A1  
Cyl No = 34  
Expiration Date = 04/05/2022  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.14.00 09/16  
06/04/2020      13:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:09
02 Std. Gas	0.081	13:09
03 Room Air	0.000	13:09
04 Std. Gas	0.082	13:10
05 Room Air	0.000	13:10
06 Std. Gas	0.081	13:11
07 Room Air	0.000	13:11

Lot No = 05620080A1  
Cyl No = 34  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.14.00 09/16  
06/04/2020      13:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:12
02 Std. Gas	0.082	13:12
03 Room Air	0.000	13:13
04 Std. Gas	0.082	13:13
05 Room Air	0.000	13:14
06 Std. Gas	0.082	13:14
07 Room Air	0.000	13:15

Lot No = 05620080A1  
Cyl No = 34  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.14.00 09/16  
06/04/2020      13:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:15
02 Std. Gas	0.082	13:16
03 Room Air	0.000	13:16
04 Std. Gas	0.082	13:17
05 Room Air	0.000	13:17
06 Std. Gas	0.081	13:18
07 Room Air	0.000	13:18

Lot No = 05620080A1  
Cyl No = 34  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000