

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006504 Location: ToxL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 994 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3064 Lot #: 20070 Exp. Date: 2-13-22
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3057 Lot #: 19010 Exp. Date: 1-3-21

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: IC56 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 056020080A1 Cyl #: 34 Exp. Date: 4-5-22

| | | |
|----------------------|---------------------|---------------------|
| Test 1 <u>0.081</u> | Test 4 <u>0.081</u> | Test 7 <u>0.081</u> |
| Test 2 <u>0.082</u> | Test 5 <u>0.080</u> | Test 8 <u>0.081</u> |
| Test 3 <u>0.081</u> | Test 6 <u>0.081</u> | Test 9 <u>0.080</u> |
| Average <u>0.081</u> | | |

C. Remarks/Maintenance: ANNUAL INSPECTION

GAS CYL LOT # 05620080A1

Instrument is acceptable to be used in the field.

Breath Analyst Signature

Reviewed by

Date

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
06/05/2020 11:34

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 12
Standard Expiration? 11/05/2021
Oper No? 666666

Flow Cal. Date: 08/20/2015
Slope 681
Intercept -673151

IR Calibration Date: 05/31/2017

| | 3um | 9um |
|--------------------|--------|--------|
| 0th Coef(*100): | -21488 | -16160 |
| 1st Coef(*100): | 265581 | 137324 |
| 2nd Coef(*100): | 2507 | 1149 |
| H2O adj(mg/l*10k): | 519 | 348 |

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
06/05/2020 11:39

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 11:40 |
| 02 Std. Sol. | 0.020 | 11:40 |
| 03 Room Air | 0.000 | 11:41 |
| 04 Std. Sol. | 0.021 | 11:42 |
| 05 Room Air | 0.000 | 11:42 |
| 06 Std. Sol. | 0.021 | 11:43 |
| 07 Room Air | 0.000 | 11:43 |

08 Sim Temp = 34.0°C

Simul Ser No = MP3064
Std Sol No = 20070
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
06/05/2020 11:44

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 11:45 |
| 02 Std. Sol. | 0.296 | 11:46 |
| 03 Room Air | 0.000 | 11:46 |
| 04 Std. Sol. | 0.298 | 11:47 |
| 05 Room Air | 0.000 | 11:48 |
| 06 Std. Sol. | 0.299 | 11:48 |
| 07 Room Air | 0.000 | 11:49 |

08 Sim Temp = 34.0°C

Simul Ser No = MP3057
Std Sol No = 19010
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
06/05/2020 11:51

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 11:51 |
| 02 *Subject Test | INT* | 11:52 |
| 03 Room Air | 0.000 | 11:52 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 34
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
06/05/2020 11:53

| Test | AC | Time |
|------------------|-------|-------|
| 01 Diagnostic | OK | 11:54 |
| 02 Room Air | 0.000 | 11:54 |
| 03 *Subject Test | RFI* | 11:54 |
| 04 Room Air | 0.000 | 11:55 |

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 34
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
06/05/2020 11:55

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 11:55 |
| 02 Std. Gas | 0.081 | 11:56 |
| 03 Room Air | 0.000 | 11:56 |
| 04 Std. Gas | 0.082 | 11:57 |
| 05 Room Air | 0.000 | 11:57 |
| 06 Std. Gas | 0.081 | 11:58 |
| 07 Room Air | 0.000 | 11:58 |

Lot No = 05620080A1
Cyl No = 34
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
06/05/2020 11:58

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 11:59 |
| 02 Std. Gas | 0.081 | 11:59 |
| 03 Room Air | 0.000 | 12:00 |
| 04 Std. Gas | 0.080 | 12:00 |
| 05 Room Air | 0.000 | 12:01 |
| 06 Std. Gas | 0.081 | 12:01 |
| 07 Room Air | 0.000 | 12:02 |

Lot No = 05620080A1
Cyl No = 34
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
06/05/2020 12:02

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 12:02 |
| 02 Std. Gas | 0.081 | 12:03 |
| 03 Room Air | 0.000 | 12:03 |
| 04 Std. Gas | 0.081 | 12:04 |
| 05 Room Air | 0.000 | 12:04 |
| 06 Std. Gas | 0.080 | 12:04 |
| 07 Room Air | 0.000 | 12:05 |

Lot No = 05620080A1
Cyl No = 34
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Form 106-I8000