

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005950

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 1011 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP3061 Lot #: 201810D Exp. Date: 10.29.20
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3069 Lot #: 201902D Exp. Date: 2.20.21

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 05620080A1 Cyl #: 34 Exp. Date: 4.5.22

Test 1	<u>0.082</u>	Test 4	<u>0.083</u>	Test 7	<u>0.083</u>
Test 2	<u>0.082</u>	Test 5	<u>0.082</u>	Test 8	<u>0.082</u>
Test 3	<u>0.082</u>	Test 6	<u>0.083</u>	Test 9	<u>0.082</u>
Average	<u>0.082</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Breath Analyst Signature

Reviewed by

Date

Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005950  
Location = TOXL                      8164.13.00 06/09  
06/10/2020                              10:17

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

Current Instrument Setup


Data Entry Mode:                      Enabled  
Start Test Sequence:                  DABACABA  
Display Prelim Rslt?                  Yes  
Display Third Digit?                  Yes  
Inhib Printer(Y/N)?                  No  
Display Volume?                        No  
Disable On Memfull?                  Yes  
# of Print Copies?                     1  
Select Std (D/W/I)?                  Dry  
Standard Value?                        0.080  
Standard Lot #?                        05620080A1  
Standard Cyl #?                        34  
Standard Expiration?                  04/05/2022  
Oper No?                                666666

Flow Cal. Date:                        06/20/2013  
Slope                                    713  
Intercept                                -723515

IR Calibration Date:                  06/20/2013  
    3um                                      9um

-----  
0th Coef(\*100):                        -37081                                  -23816  
1st Coef(\*100):                        290549                                  133411  
2nd Coef(\*100):                        552                                        1090  
H2O adj(mg/l\*10k):                    683                                        502

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = TOXL      8164.13.00 06/09  
06/10/2020      10:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:35
02 Std. Sol.	0.017	10:35
03 Room Air	0.000	10:36
04 Std. Sol.	0.017	10:37
05 Room Air	0.000	10:37
06 Std. Sol.	0.018	10:38
07 Room Air	0.000	10:38

08 Sim Temp = 34.0°C

Simul Ser No = MP3061

Std Sol No = 201810D

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = TOXL      8164.13.00 06/09  
06/10/2020      10:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:44
02 Std. Sol.	0.295	10:45
03 Room Air	0.000	10:46
04 Std. Sol.	0.294	10:46
05 Room Air	0.000	10:47
06 Std. Sol.	0.295	10:48
07 Room Air	0.000	10:48

08 Sim Temp = 34.0°C

Simul Ser No = MP3069

Std Sol No = 201902D

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005950  
Location = TOXL                    8164.13.00 06/09  
06/10/2020                            10:49

Test	AC	Time
01 Room Air	0.000	10:50
02 *Subject Test	INT*	10:50
03 Room Air	0.000	10:51

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS921456  
Lot No = 05620080A1  
Cyl No = 34  
Expiration Date = 04/05/2022  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005950  
Location = TOXL                      8164.13.00 06/09  
06/10/2020                              10:51

Test	AC	Time
01 Diagnostic	OK	10:52
02 Room Air	0.000	10:52
03 *Subject Test	RFI*	10:52
04 Room Air	0.000	10:53

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                      Weight = 150  
Test = DUI                              Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 05620080A1  
Cyl No = 34  
Expiration Date = 04/05/2022  
County = 08                              Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = TOXL      8164.13.00 06/09  
06/10/2020      10:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:54
02 Std. Gas	0.082	10:54
03 Room Air	0.000	10:54
04 Std. Gas	0.082	10:55
05 Room Air	0.000	10:55
06 Std. Gas	0.082	10:56
07 Room Air	0.000	10:56

Lot No = 05620080A1  
Cyl No = 34  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-18000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = TOXL      8164.13.00 06/09  
06/10/2020      10:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:57
02 Std. Gas	0.083	10:57
03 Room Air	0.000	10:58
04 Std. Gas	0.082	10:58
05 Room Air	0.000	10:59
06 Std. Gas	0.083	10:59
07 Room Air	0.000	10:59

Lot No = 05620080A1  
Cyl No = 34  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080AC

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = TOXL      8164.13.00 06/09  
06/10/2020      11:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Gas	0.083	11:01
03 Room Air	0.000	11:01
04 Std. Gas	0.082	11:01
05 Room Air	0.000	11:02
06 Std. Gas	0.082	11:02
07 Room Air	0.000	11:03

Lot No = 05620080A1  
Cyl No = 34  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000