

INTOXILYZER® 8000 CALIBRATION ADJUSTMENT

Intoxilyzer® 8000 Serial Number: 80-004938 Location: TOXL

- A. Flow Sensor Calibration and Verification Check (Level 3,M,C,F)
1. Replaced o-rings if damaged ADJUST VERIFY
 2. Flow Meter Serial Number: 40655 & 55260
 3. Air Supplied to Intoxilyzer® 8000 at:
 - a. 5 L/min 15 L/min 30 L/min
 4. Flow Rate Calibration Printout Attached
 - a. Correlation ≥ 0.99000
 5. Flow Sensor Calibration Verification (Level 3,D,F)
 - a. 10 L/min: 0. 160 L/S X 60 Sec/min = 9.60 L/min
 - b. 20 L/min: 0. 324 L/S X 60 Sec/min = 19.44 L/min
 - c. Flow Rates within ± 1 L/min of Expected Value

- B. Gas Tank Sensor Check (Level 3,D,G)
1. Display: 933 psi Regulator: 950 psi
 2. Display and Regulator within 50 psi
 3. Completed tare of tank sensor if needed (Level 3,M,C,G)

- C. Optical Bench Calibration and Verification Check (Level 3,M,C,O)
1. Autocalibration Printout Attached
 - a. Max Power Res Value ≥ 10
 - b. Auto Range Res Value ≥ 4
 2. Simulator Solutions for Optical Bench Calibration Adjustment
 - a. Set # Solutions to Run at 5

Soln.	g/210 L	Lot No.	Exp. Date	Simulator SN
1	0.000 (ACTUAL)	NA - MilliQ H ₂ O	NA - MilliQ H ₂ O	DR7111
2	0.040 (0.040)	201808D	8.22.20	DR7347
3	0.080 (0.081)	201807C	7.25.20	DR5114
4	0.150 (0.151)	201811E	11.26.20	DR5131
5	0.300 (0.298)	19010	1.3.21	DR7346

3. 0.100 AC Calibration Gas for H₂O Adjustment
 - a. Lot No. 13518100A3 Cyl No. 4 Exp. Date: 8.5.2020
4. Atmospheric Pressure
 - a. 923 mbar Displayed by Intoxilyzer® 8000
 - b. 952 mbar Adjusted to using barometer
 - c. 952 mbar on Auto Calibration Report printout
5. Screen displayed "Calibration Success"

6. Calibration Adjustment Printout Attached
- a. Solution 1 Avg % Abs \leq 0.2500
- b. Solution 2-5 REL STD DEV \leq 3.000
- c. Residual (g/210 L) Values for Solutions 1-5 \leq 0.0020 for 3 μ m and 9 μ m channels
- d. Dry Gas H2O Adjustment Sum for 3 μ m and 9 μ m channels within \pm 10

	Average		H ₂ O Adjust	=	
3 μ m	4406	+	355	=	4761
9 μ m	4391	+	370	=	4761

7. Optical Bench Calibration Verification (Level 1, S and C)
- a. Wet Calibration Check
- i. Low AC Known Value \leq 0.03 AC: 0.020 AC
 Sim. SN: MP3064 Lot No.: 20070 Exp. Date: 2.13.22
- ii. High AC Known Value \geq 0.25 AC: 0.300^{CEE} AC 0.250 AC
 Sim. SN: MP3067 Lot No.: 201911B Exp. Date: 11.5.21
- b. Dry Calibration Check: Known Value 0.08 AC
 Lot No. 24119080A1 Cyl No. 9 Exp. Date: 11.5.21
- | | | |
|-------------------------|------------------------|------------------------|
| Test 1 <u>0.081</u> AC | Test 4 <u>0.079</u> AC | Test 7 <u>0.079</u> AC |
| Test 2 <u>0.081</u> AC | Test 5 <u>0.080</u> AC | Test 8 <u>0.079</u> AC |
| Test 3 <u>0.080</u> AC | Test 6 <u>0.079</u> AC | Test 9 <u>0.079</u> AC |
| Average <u>0.080</u> AC | | |
- c. Wet Calibration Check and Dry Calibration Check AC results are within \pm 0.005 or \pm 5% (whichever is greater) of stated value.

D. Remarks/Maintenance: CALIBRATION ADJUST DUE TO
ATMOSPHERIC MONITOR READING 923 mbar WHEN ACTUAL
ATMOSPHERIC PRESSURE IS 952 mbar.
TAPED FRAYING/TARE TEAR IN BREATH HOSE PROTECTOR.
CEE

Instrument is acceptable to be used in the field.

Charles E. Ede
 Breath Analyst Signature

6.10.20
 Date

NA
 Reviewed by

NA
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
06/10/2020 11:39

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
 SQRT(Diff) = 6.082
2: Rate (Liters/min) = 15
 SQRT(Diff) = 11.488
3: Rate (Liters/min) = 30
 SQRT(Diff) = 22.469


Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 587

Rounded Intercept = -338327

Correlation = 0.99688



TOXL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004938
 06/10/2020 12:29:04

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>>			<<<<< 9um >>>>>		

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.1100	(0.0100)	0.1500	(0.0070)	
Sample #2	0.0860	(0.0990)	0.1610	(0.0330)	
Sample #3	0.0740	(0.1540)	0.1670	(0.0430)	
Sample #4	0.0620	(0.1970)	0.1590	(0.0590)	
Avg % Abs	0.0740	(0.1500)	0.1623	(0.0450)	
STD DEV	0.0120	(0.0491)	0.0042	(0.0131)	
REL STD DEV	16.216	(32.748)	2.565	(29.144)	

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.7920	(-0.0030)	1.5710	(0.0180)	
Sample #2	0.7930	(0.0190)	1.5560	(0.0400)	
Sample #3	0.8040	(0.0400)	1.5550	(0.0510)	
Sample #4	0.7960	(0.0560)	1.5720	(0.0500)	
Avg % Abs	0.7977	(0.0383)	1.5610	(0.0470)	
STD DEV	0.0057	(0.0186)	0.0095	(0.0061)	
REL STD DEV	0.713	(48.408)	0.611	(12.942)	

Solution = 0.081 g/210L or 0.3857 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	1.5080	(-0.0080)	2.9150	(0.0000)	
Sample #2	1.5160	(0.0160)	2.9240	(-0.0010)	
Sample #3	1.5140	(0.0150)	2.9440	(0.0000)	
Sample #4	1.5040	(0.0280)	2.9250	(0.0000)	
Avg % Abs	1.5113	(0.0197)	2.9310	(-0.0003)	
STD DEV	0.0064	(0.0072)	0.0113	(0.0006)	
REL STD DEV	0.425	(36.784)	0.384	(173.205)	

Solution = 0.151 g/210L or 0.7190 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	2.6980	(0.0040)	5.2770	(0.0000)	
Sample #2	2.7250	(0.0080)	5.2770	(0.0230)	
Sample #3	2.7050	(0.0130)	5.2420	(0.0420)	
Sample #4	2.7020	(0.0280)	5.2430	(0.0460)	
Avg % Abs	2.7107	(0.0163)	5.2540	(0.0370)	
STD DEV	0.0125	(0.0104)	0.0199	(0.0123)	
REL STD DEV	0.461	(63.724)	0.379	(33.211)	

Solution = 0.298 g/210L or 1.4190 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	5.2080	(0.0000)	9.9580	(0.0110)	
Sample #2	5.2430	(0.0120)	9.9370	(0.0550)	
Sample #3	5.2440	(0.0240)	9.9460	(0.0560)	
Sample #4	5.2500	(0.0150)	9.9300	(0.0620)	
Avg % Abs	5.2457	(0.0170)	9.9377	(0.0577)	
STD DEV	0.0038	(0.0062)	0.0080	(0.0038)	
REL STD DEV	0.072	(36.735)	0.081	(6.565)	

TOXL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004938
 06/10/2020 12:29:04

Auto Calibration

pg 2 of 2

<<<<< 3um >>>>>

 Zero Order Coef -224.34
 First Order Coef 2698.49
 Second Order Coef 9.63

<<<<< 9um >>>>>

 -231.79
 1367.90
 8.40

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.001	0.0005
0.040	0.041	-0.0006
0.081	0.081	-0.0004
0.151	0.150	0.0006
0.298	0.298	-0.0001

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0002
0.040	0.040	-0.0004
0.081	0.081	0.0002
0.151	0.151	0.0001
0.298	0.298	-0.0000


<<<<< 3um >>>>> <<<<< 9um >>>>>

 Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1
 Sample

Sample #1	4422.00	4405.00
Sample #2	4370.00	4395.00
Sample #3	4405.00	4383.00
Sample #4	4445.00	4395.00
Avg	4406.6665	4391.0000
STD DEV	37.5278	6.9282
REL STD DEV	0.852	0.158
H2O adjust (mg/l*10k)	355	370

Atmospheric Pressure = 952

*****CALIBRATION SUCCESSFUL*****



TOXL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004938
 06/10/2020 12:29:04

Auto Calibration
 Max Power Res Value = 41
 Auto Range Res Value = 20

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
06/10/2020 13:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Sol.	0.020	13:12
03 Room Air	0.000	13:13
04 Std. Sol.	0.020	13:13
05 Room Air	0.000	13:14
06 Std. Sol.	0.020	13:15
07 Room Air	0.000	13:15

08 Sim Temp = 34.0°C

Simul Ser No = MP3064
Std Sol No = 20070
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
06/10/2020 13:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:18
02 Std. Sol.	0.254	13:18
03 Room Air	0.000	13:19
04 Std. Sol.	0.253	13:20
05 Room Air	0.000	13:20
06 Std. Sol.	0.253	13:21
07 Room Air	0.000	13:22

08 Sim Temp = 34.0°C

Simul Ser No = MP3067
Std Sol No = 201911B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.250 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
06/10/2020 13:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Gas	0.081	13:29
03 Room Air	0.000	13:30
04 Std. Gas	0.081	13:30
05 Room Air	0.000	13:30
06 Std. Gas	0.080	13:31
07 Room Air	0.000	13:31

Lot No = 24119080A1
Cyl No = 9
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

⁷
CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
06/10/2020 13:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:32
02 Std. Gas	0.079	13:32
03 Room Air	0.000	13:33
04 Std. Gas	0.080	13:33
05 Room Air	0.000	13:34
06 Std. Gas	0.079	13:34
07 Room Air	0.000	13:35

Lot No = 24119080A1
Cyl No = 9
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
06/10/2020 13:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:37
02 Std. Gas	0.079	13:37
03 Room Air	0.000	13:37
04 Std. Gas	0.079	13:38
05 Room Air	0.000	13:38
06 Std. Gas	0.079	13:39
07 Room Air	0.000	13:39

Lot No = 24119080A1
Cyl No = 9
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Form 106-I8000