



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Gabriel Johnston</i>		
Location <i>Siqueler County (Langdon)</i>	Intoxilyzer® Serial Number <i>80-005951</i>	
Gas Lot Number <i>24119080A1</i>	Gas Cylinder Number <i>1</i>	Gas Expiration Date <i>11-05-2021</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.081* AC
 B. *0.080* AC
 C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>6-15-2020</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>6-18-2020</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005951
Location = LANG 8164.14.00 09/16
06/15/2020 08:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:37
02 Std. Gas	0.081	08:37
03 Room Air	0.000	08:38
04 Std. Gas	0.080	08:38
05 Room Air	0.000	08:39
06 Std. Gas	0.079	08:39
07 Room Air	0.000	08:39

Lot No = 24119080A1
Cyl No = 1
Exp Date = 11/05/2021
County = 10 Oper No. = 021009



Operator Signature
GABRIEL JOHNSTON

Remarks:

Form 106-I8000