

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006499

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 679 psi Regulator: 700 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq$  0.03 AC: 0.030 AC  
Sim. Ser #: MP3063 Lot #: 20130 Exp. Date: 3.10.22
- b.  High AC (Level 1,C): Known Value  $\geq$  0.25 AC: 0.300 AC  
Sim. Ser #: MP3069 Lot #: 201902D Exp. Date: 2.20.21

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS 6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 05620080A1 Cyl #: 34 Exp. Date: 4.5.22

Test 1	<u>0.080</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.081</u>	Test 8	<u>0.082</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.082</u>
Average	<u>0.081</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION  
REPLACED BREATH HOSE O-RING

Instrument is acceptable to be used in the field.

Charles E. Ehr  
Breath Analyst Signature

6.30.2020  
Date

NA  
Reviewed by

NA  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006499  
Location = TOXL                    8164.14.00 09/16  
06/30/2020                            14:34

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            05620080A1  
Standard Cyl #?            34  
Standard Expiration?       04/05/2022  
Oper No?                    666666

Flow Cal. Date:            08/18/2015  
Slope                        678  
Intercept                   -685002

IR Calibration Date:       08/18/2015  
                                  3um                            9um

-----  
0th Coef(\*100):            -22123                        -16014  
1st Coef(\*100):            264933                        138505  
2nd Coef(\*100):            3281                            1621  
H2O adj(mg/l\*10k):        704                            456

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006499  
Location = TOXL      8164.14.00 09/16  
06/30/2020      14:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Sol.	0.029	14:37
03 Room Air	0.000	14:38
04 Std. Sol.	0.029	14:39
05 Room Air	0.000	14:39
06 Std. Sol.	0.030	14:40
07 Room Air	0.000	14:41

08 Sim Temp = 34.0°C

Simul Ser No = MP3063

Std Sol No = 20130

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.030 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006499  
Location = TOXL      8164.14.00 09/16  
06/30/2020      14:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:42
02 Std. Sol.	0.300	14:43
03 Room Air	0.000	14:44
04 Std. Sol.	0.303	14:45
05 Room Air	0.000	14:45
06 Std. Sol.	0.303	14:46
07 Room Air	0.000	14:47


08 Sim Temp = 34.0°C

Simul Ser No = MP3069

Std Sol No = 201902D

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006499  
Location = TOXL      8164.14.00 09/16  
06/30/2020      14:48

Test	AC	Time
01 Room Air	0.000	14:49
02 *Subject Test	INT*	14:49
03 Room Air	0.000	14:50

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS921456  
Lot No = 05620080A1  
Cyl No = 34  
Expiration Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006499  
Location = TOXL                 8164.14.00 09/16  
06/30/2020                        14:50

Test	AC	Time
01 Diagnostic	OK	14:51
02 Room Air	0.000	14:51
03 *Subject Test	RFI*	14:52
04 Room Air	0.000	14:53

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 05620080A1  
Cyl No = 34  
Expiration Date = 04/05/2022  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006499  
Location = TOXL      8164.14.00 09/16  
06/30/2020      14:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:54
02 Std. Gas	0.080	14:54
03 Room Air	0.000	14:55
04 Std. Gas	0.081	14:55
05 Room Air	0.000	14:56
06 Std. Gas	0.081	14:56
07 Room Air	0.000	14:57

Lot No = 05620080A1  
Cyl No = 34  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006499  
Location = TOXL      8164.14.00 09/16  
06/30/2020      14:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:57
02 Std. Gas	0.081	14:58
03 Room Air	0.000	14:58
04 Std. Gas	0.081	14:59
05 Room Air	0.000	14:59
06 Std. Gas	0.081	15:00
07 Room Air	0.000	15:00

Lot No = 05620080A1  
Cyl No = 34  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006499  
Location = TOXL      8164.14.00 09/16  
06/30/2020      15:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:01
02 Std. Gas	0.081	15:01
03 Room Air	0.000	15:02
04 Std. Gas	0.082	15:02
05 Room Air	0.000	15:03
06 Std. Gas	0.082	15:03
07 Room Air	0.000	15:04


Lot No = 05620080A1

Cyl No = 34

Exp Date = 04/05/2022

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000