

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6502

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 673 psi Regulator: 700 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3064 Lot #: 20070 Exp. Date: 2.13.22
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP5319 Lot #: 20201B Exp. Date: 1.28.22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 2411908041 Cyl #: 9 Exp. Date: 11.5.2021

| | | | | | |
|---------|--------------|--------|--------------|--------|--------------|
| Test 1 | <u>0.081</u> | Test 4 | <u>0.082</u> | Test 7 | <u>0.082</u> |
| Test 2 | <u>0.082</u> | Test 5 | <u>0.082</u> | Test 8 | <u>0.082</u> |
| Test 3 | <u>0.082</u> | Test 6 | <u>0.082</u> | Test 9 | <u>0.082</u> |
| Average | <u>0.082</u> | | | | |

C. Remarks/Maintenance: ANNUAL INSPECTION

NEW EXTERNAL BATTERY

Instrument is acceptable to be used in the field.

Charles E. Egan
Breath Analyst Signature

NA

6.30.2020
Date

NA

Reviewed by

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006502
Location = TOXL 8164.14.00 09/16
06/30/2020 14:29

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^ &* () _ + ?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^ &* () _ + ?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 9
Standard Expiration? 11/05/2021
Oper No? 666666

Flow Cal. Date: 08/19/2015
Slope 674
Intercept -654491

IR Calibration Date: 08/19/2015

| | 3um | 9um |
|--------------------|--------|--------|
| 0th Coef(*100): | -24446 | -18103 |
| 1st Coef(*100): | 267787 | 135417 |
| 2nd Coef(*100): | 3387 | 1610 |
| H2O adj(mg/l*10k): | 740 | 513 |

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006502
Location = TOXL 8164.14.00 09/16
06/30/2020 14:33

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 14:33 |
| 02 Std. Sol. | 0.018 | 14:34 |
| 03 Room Air | 0.000 | 14:35 |
| 04 Std. Sol. | 0.018 | 14:35 |
| 05 Room Air | 0.000 | 14:36 |
| 06 Std. Sol. | 0.018 | 14:37 |
| 07 Room Air | 0.000 | 14:37 |

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 20070

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

LOW AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006502
Location = TOXL 8164.14.00 09/16
06/30/2020 14:44

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 14:44 |
| 02 Std. Sol. | 0.294 | 14:45 |
| 03 Room Air | 0.000 | 14:46 |
| 04 Std. Sol. | 0.295 | 14:47 |
| 05 Room Air | 0.000 | 14:47 |
| 06 Std. Sol. | 0.294 | 14:48 |
| 07 Room Air | 0.000 | 14:48 |

08 Sim Temp = 34.0°C

Simul Ser No = MP5319
Std Sol No = 202001B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006502
Location = TOXL 8164.14.00 09/16
06/30/2020 14:50

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 14:50 |
| 02 *Subject Test | INT* | 14:51 |
| 03 Room Air | 0.000 | 14:52 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006502
Location = TOXL 8164.14.00 09/16
06/30/2020 14:54

| Test | AC | Time |
|------------------|-------|-------|
| 01 Diagnostic | OK | 14:54 |
| 02 Room Air | 0.000 | 14:55 |
| 03 *Subject Test | RFI* | 14:55 |
| 04 Room Air | 0.000 | 14:55 |

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006502
Location = TOXL 8164.14.00 09/16
06/30/2020 14:59

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 15:00 |
| 02 Std. Gas | 0.081 | 15:00 |
| 03 Room Air | 0.000 | 15:00 |
| 04 Std. Gas | 0.082 | 15:01 |
| 05 Room Air | 0.000 | 15:01 |
| 06 Std. Gas | 0.082 | 15:02 |
| 07 Room Air | 0.000 | 15:02 |

Lot No = 24119080A1
Cyl No = 9
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006502
Location = TOXL 8164.14.00 09/16
06/30/2020 15:02

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 15:03 |
| 02 Std. Gas | 0.082 | 15:03 |
| 03 Room Air | 0.000 | 15:04 |
| 04 Std. Gas | 0.082 | 15:04 |
| 05 Room Air | 0.000 | 15:05 |
| 06 Std. Gas | 0.082 | 15:05 |
| 07 Room Air | 0.000 | 15:06 |

Lot No = 24119080A1
Cyl No = 9
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006502
Location = TOXL 8164.14.00 09/16
06/30/2020 15:06

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 15:06 |
| 02 Std. Gas | 0.082 | 15:07 |
| 03 Room Air | 0.000 | 15:07 |
| 04 Std. Gas | 0.082 | 15:08 |
| 05 Room Air | 0.000 | 15:08 |
| 06 Std. Gas | 0.082 | 15:08 |
| 07 Room Air | 0.000 | 15:09 |

Lot No = 24119080A1
Cyl No = 9
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000