




INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

| | |
|---|---|
| Serial Number 80-006491 | Instrument Location Langdon County Sheriff's Office (Langdon) |
| Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____ | |


Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

| | |
|---|-------------------------|
| Field Inspector Signature  4372 | Date 7-2-2020 |
|---|-------------------------|

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

| | |
|--|------------------------------------|
| Reviewed/Certified By  | Certified Date 7-14-2020 |
|--|------------------------------------|

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = LAMO 8164.14.00 09/16
07/02/2020 18:15

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05620080A1
Standard Cyl #? 12
Standard Expiration? 04/05/2022
Oper No? 132539

Flow Cal. Date: 05/22/2018
Slope 639
Intercept -618311

IR Calibration Date: 05/23/2018

| | 3um | 9um |
|--------------------|--------|--------|
| 0th Coef(*100): | -23641 | -17066 |
| 1st Coef(*100): | 264855 | 137703 |
| 2nd Coef(*100): | 2663 | 1471 |
| H2O adj(mg/l*10k): | 556 | 451 |

***** Printer Test End *****


Operator Signature
JOSHUA FASTNACHT

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = LAMO 8164.14.00 09/16
07/02/2020 18:17

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 18:17 |
| 02 Std. Gas | 0.078 | 18:18 |
| 03 Room Air | 0.000 | 18:18 |
| 04 Std. Gas | 0.077 | 18:19 |
| 05 Room Air | 0.000 | 18:19 |
| 06 Std. Gas | 0.078 | 18:20 |
| 07 Room Air | 0.000 | 18:20 |

Lot No = 05620080A1
Cyl No = 12
Exp Date = 04/05/2022
County = 23 Oper No. = 132539


Operator Signature
JOSHUA FASTNACHT

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = LAMO 8164.14.00 09/16
07/02/2020 18:22

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 18:22 |
| 02 Std. Gas | 0.077 | 18:23 |
| 03 Room Air | RFI* | 18:23 |
| 04 Room Air | 0.000 | 18:24 |

*Invalid Test
Inhibited - RFI

Lot No = 05620080A1
Cyl No = 12
Exp Date = 04/05/2022
County = 23 Oper No. = 132539


Operator Signature
JOSHUA FASTNACHT

Remarks:

Form 106-I8000