

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006502
Location = STUT 8164.14.00 09/16
07/12/2020 13:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:27
02 Std. Gas	0.081	13:27
03 Room Air	RFI*	13:28
04 Room Air	0.000	13:28

*Invalid Test
Inhibited - RFI

Lot No = 05620080A1
Cyl No = 4
Exp Date = 04/05/2022
County = 47 Oper No. = 100012



Operator Signature
DANA RUSTEBAKKE

Remarks:

RFI Test - Good

Form 106-I8000