



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-006490	Instrument Location ASHL
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Bryan Thomas</i>	Date 10/15/2020
--	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Charles E. Ehl</i>	Certified Date 10/19/2020
--	-------------------------------------

*CEE 10/19/2020
 CALLED MCINTOSH COUNTY S.O. TO HAVE THEM UPDATE
 LOCATION CODE FROM TOYL TO ASHL, LOCATION CODE
 DOES NOT EFFECT INSTALL OR INSTRUMENT FUNCTION.*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
10/15/2020 22:13

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 28
Standard Expiration? 11/05/2021
Oper No? 132584

Flow Cal. Date: 09/04/2020
Slope 649
Intercept -598817

IR Calibration Date: 09/04/2020
 3um 9um

0th Coef(*100):	-25190	-16392
1st Coef(*100):	271645	136906
2nd Coef(*100):	2533	1222
H2O adj(mg/l*10k):	460	309

***** Printer Test End *****



Operator Signature
DYLAN THOMES

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
10/15/2020 22:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:15
02 Std. Gas	0.080	22:15
03 Room Air	0.000	22:16
04 Std. Gas	0.080	22:16
05 Room Air	0.000	22:17
06 Std. Gas	0.080	22:17
07 Room Air	0.000	22:18

Lot No = 24119080A1
Cyl No = 28
Exp Date = 11/05/2021
County = 26 Oper No. = 132584


Operator Signature
DYLAN THOMES

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
10/15/2020 22:20

Test	AC	Time
01 Room Air	0.000	22:22
02 *Subject Test	RFI*	22:22
03 Room Air	0.000	22:23

*Invalid Test
Inhibited - RFI

Sub Name = DOE, JOHN ABC
Sub DOB = 01/01/2001
Sub Sex = Male Weight = 300
Test = DUI Cit = NA
Dr. Lic. = ND/DOE123456
Lot No = 24119080A1
Cyl No = 28
Expiration Date = 11/05/2021
County = 26 Oper No. = 132584



Operator Signature
DYLAN THOMES

Remarks:

Form 106-I8000