



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Alexander Robert Wehrung</i>		
Location <i>NAPO</i>	Intoxilyzer® Serial Number <i>80-005948</i>	
Gas Lot Number <i>05G20080A1</i>	Gas Cylinder Number <i>46</i>	Gas Expiration Date <i>04/05/2022</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.080* AC
 - B. *0.080* AC
 - C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>10/28/2020</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>10.28.2020</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005948
Location = NAPO 8164.14.00 09/16
10/28/2020 09:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:53
02 Std. Gas	0.080	09:53
03 Room Air	0.000	09:54
04 Std. Gas	0.080	09:54
05 Room Air	0.000	09:55
06 Std. Gas	0.080	09:55
07 Room Air	0.000	09:56

Lot No = 05620080A1
Cyl No = 46
Exp Date = 04/05/2022
County = 24 Oper No. = 132466


Operator Signature
ALEXANDER WEHRUNG

Remarks: *ACA test*

Form 106-I8000