



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Anthony G. Dumas SR.</i>		
Location <i>Grafton (GFTN)</i>	Intoxilyzer® Serial Number <i>80-006501</i>	
Gas Lot Number <i>05620080 A1</i>	Gas Cylinder Number <i>004</i>	Gas Expiration Date <i>4/5/2022</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.078* AC
 B. *0.078* AC
 C. *0.078* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency. *N/A*
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>02 Nov 2020</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>11/3/2020</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = GFTN 8164.14.00 09/16
11/02/2020 15:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:03
02 Std. Gas	0.078	15:03
03 Room Air	0.000	15:04
04 Std. Gas	0.078	15:04
05 Room Air	0.000	15:05
06 Std. Gas	0.078	15:05
07 Room Air	0.000	15:05

Lot No = 05620080A1
Cyl No = 6
Exp Date = 04/05/2022
County = 50 Oper No. = 020305



Operator Signature
ANTHONY DUMAS

Remarks:

Form 106-I8000