



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Kathryn M. Allen</i>		
Location <i>Cass County Jail</i>	Intoxilyzer® Serial Number <i>80-006496</i>	
Gas Lot Number <i>05620080A1</i>	Gas Cylinder Number <i>27</i>	Gas Expiration Date <i>4-5-2022</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0. *081* AC
 B. 0. *081* AC
 C. 0. *081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature <i>Kathryn M. Allen</i>	Date <i>1-12-2021</i>
Reviewed By (Crime Laboratory Use Only) <i>Paul E. E...</i>	Date <i>1.13.2021</i>

