




**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-005358</b>	Instrument Location <b>BMDC</b>
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Monthly Tests</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature  4850	Date <b>1-28-21</b>
-----------------------------------------------------------------------------------------------------------------------	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date <b>2.3.21</b>
--------------------------------------------------------------------------------------------------------------	---------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005358  
Location = BMDC                8164.14.00 09/16  
01/28/2021                      16:53

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            24119080A1  
Standard Cyl #?            30  
Standard Expiration?       11/05/2021  
Oper No?                    132507

Flow Cal. Date:            06/04/2019  
Slope                      570  
Intercept                  4079

IR Calibration Date:       06/04/2019  
                                    3um                      9um

-----  
0th Coef(\*100):            -19471                      -24652  
1st Coef(\*100):            273237                      143582  
2nd Coef(\*100):            2965                        1293  
H2O adj(mg/l\*10k):        273                         374

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
-----  
Operator Signature  
KYLE HAMAN

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005358  
Location = BMDC      8164.14.00 09/16  
01/28/2021      16:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:54
02 Std. Gas	0.082	16:55
03 Room Air	0.000	16:55
04 Std. Gas	0.082	16:56
05 Room Air	0.000	16:56
06 Std. Gas	0.082	16:57
07 Room Air	0.000	16:57

Lot No = 24119080A1  
Cyl No = 30  
Exp Date = 11/05/2021  
County = 08      Oper No. = 132507

  
Operator Signature  
KYLE HAMAN

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005358  
Location = BMDC                8164.14.00 09/16  
01/28/2021                        16:59

Test	AC	Time
01 Room Air	RFI*	17:03
02 Room Air	0.000	17:03

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI TEST, NA NA  
Sub DOB = 10/10/2010  
Sub Sex = Unknown                Weight = NA  
Test = OTH                        Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 24119080A1  
Cyl No = 30  
Expiration Date = 11/05/2021  
County = 08                        Oper No. = 132507

 4850

Operator Signature  
KYLE HAMAN

Remarks:

Form 106-I8000