



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Kari Stuart		
Location MCCF: 1201 14th St. SE, Watford City, ND 58054		Intoxilyzer® Serial Number 80-004954
Gas Lot Number 051020080 A1	Gas Cylinder Number 20	Gas Expiration Date 4/5/2022

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0.001 AC
 B. 0.001 AC
 C. 0.002 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature Kari Stuart #3416	Date 2/2/21
Reviewed By (Crime Laboratory Use Only) [Signature]	Date 2/8/21


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = MCSO 8164.14.00 09/16
02/02/2021 02:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	02:38
02 Std. Gas	0.081	02:38
03 Room Air	0.000	02:39
04 Std. Gas	0.081	02:39
05 Room Air	0.000	02:40
06 Std. Gas	0.082	02:40
07 Room Air	0.000	02:40

Lot No = 05620080A1
Cyl No = 28
Exp Date = 04/05/2022
County = 27 Oper No. = 132007

 #3416

Operator Signature
KARI STUART


Remarks:

Form 106-I8000

STATE OF NORTH DAKOTA }
COUNTY OF MCKENZIE }^{ss}

I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Watford City, North Dakota.

on this 14 day of February, 2021

 #3416
Mckenzie County Sheriff's Department