



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-005947	Instrument Location Dickinson LEC
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Breath hose replacement</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): Replaced breath hose
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 3/1/21
-------------------------------	-----------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date 3/4/2021
---------------------------	-----------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
03/01/2021 17:12

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 32
Standard Expiration? 03/05/2023
Oper No? 130935

Flow Cal. Date: 06/21/2013
Slope 693
Intercept -735360

IR Calibration Date: 04/06/2015
 3um 9um

0th Coef(*100): -28096 -26982
1st Coef(*100): 280034 136042
2nd Coef(*100): 3530 1567
H2O adj(mg/l*10k): 510 437

***** Printer Test End *****


Operator Signature
KEVIN ELDRIDGE

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
03/01/2021 17:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:13
02 Std. Gas	0.079	17:14
03 Room Air	0.000	17:14
04 Std. Gas	0.079	17:15
05 Room Air	0.000	17:15
06 Std. Gas	0.079	17:16
07 Room Air	0.000	17:16

Lot No = 02621080A1
Cyl No = 32
Exp Date = 03/05/2023
County = 45 Oper No. = 130935


Operator Signature
KEVIN ELDRIDGE

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
03/01/2021 17:17

Test	AC	Time
01 Room Air	RFI*	17:18
02 Room Air	RFI*	17:18

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST NMN
Sub DOB = 01/01/2000
Sub Sex = Male Weight = 175
Test = DUI Cit = 9999999
Dr. Lic. = ND/NONE
Lot No = 02621080A1
Cyl No = 32
Expiration Date = 03/05/2023
County = 45 Oper No. = 130935


Operator Signature
KEVIN ELDRIDGE

Remarks:

Form 106-I8000