

## ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Jonathan Lampi			
Location Grand Forks Police Department		Intoxilyzer® Serial Number 80-005360	
Gas Lot Number 02621080A1	Gas Cylinder Number 35		Gas Expiration Date 03-05-23

## Check When Done:

坏	1.	Scan/Enter	Gas C	ylinder	Information	(Level	1, Function	S).
---	----	------------	-------	---------	-------------	--------	-------------	-----

2. Perform an ACA Test (Level 1, Function C).

Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

A. 0.082 AC

B. 0.083 AC C. 0.080 AC

- ☑ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test operator Signature		Date 03-02-21
Reviewed By (Crime Laboratory Use Only)	Challe El	Date 3.2.21

**FORM 105-G** 

Intoxilyzer Test Record and Checklist NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer
North Dakota Model 8000
Location = GFPD
03/02/2021

SN 80-005360 8164.14.00 09/16 12:43

## DRY CAL CHECK

Test	AC	Time	
01 Room Air	0.000	12:43	
02 Std. Gas	0.082	12:44	
03 Room Air	0.000	12:44	
04 Std. Gas	0.083	12:45	
05 Room Air	0.000	12:45	
06 Std. Gas	0.080	12:45	
07 Room Air	0.000	12:46	

Lot No = 02621080A1

Cyl No = 35

Exp Date = 03/05/2023

County = 18

Oper No. = 039947

Remarks:

Form 106-I8000

Operator Signature
JONATHAN LAMPI

gas cylmder replace.