




INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

| | |
|--|------------------------------------|
| Serial Number 80-004945 | Instrument Location BMDC |
| Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Monthly Tests</u> | |

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

| | |
|---|-------------------------|
| Field Inspector Signature  4850 | Date 3/1/2021 |
|---|-------------------------|

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

| | |
|--|-----------------------------------|
| Reviewed/Certified By  | Certified Date 3/9/2021 |
|--|-----------------------------------|


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = BMDC 8164.14.00 09/16
03/01/2021 12:18

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 12:19 |
| 02 Std. Gas | 0.082 | 12:19 |
| 03 Room Air | 0.000 | 12:20 |
| 04 Std. Gas | 0.081 | 12:20 |
| 05 Room Air | 0.000 | 12:21 |
| 06 Std. Gas | 0.082 | 12:21 |
| 07 Room Air | 0.000 | 12:21 |

Lot No = 24119080A1
Cyl No = 29
Exp Date = 11/05/2021
County = 08 Oper No. = 132507

 4850

Operator Signature
KYLE HAMAN

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = BMDC 8164.14.00 09/16
03/01/2021 12:24

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | RFI* | 12:25 |
| 02 Room Air | 0.000 | 12:26 |

*Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, NA NA
Sub DOB = 09/09/2000
Sub Sex = Unknown Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 24119080A1
Cyl No = 29
Expiration Date = 11/05/2021
County = 08 Oper No. = 132507


Operator Signature
KYLE HAMAN

Remarks:

Form 106-I8000