

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004948

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 215 psi Regulator: 250 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3064 Lot #: 20070 Exp. Date: 2.13.22
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP5317 Lot #: 202012A Exp. Date: 12.8.22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 24119080A21 Cyl #: 6 Exp. Date: 11.5.21
Test 1 0.082 Test 4 0.082 Test 7 0.082
Test 2 0.083 Test 5 0.082 Test 8 0.083
Test 3 0.083 Test 6 0.082 Test 9 0.083
Average 0.082

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED EXTERNAL BATTERY

Instrument is acceptable to be used in the field.

Charles E. Ed
Breath Analyst Signature

4.13.21
Date

Kali L. Diet
Reviewed by

4-13-2021
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.14.00 09/16
04/13/2021 16:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 6
Standard Expiration? 11/05/2021
Oper No? 666666

Flow Cal. Date: 08/28/2018
Slope 642
Intercept -606282

IR Calibration Date: 08/28/2018
 3um 9um

0th Coef(*100): -13609 -14548
1st Coef(*100): 248186 128726
2nd Coef(*100): 3538 1508
H2O adj(mg/l*10k): 602 515

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.14.00 09/16
04/13/2021 16:33

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:33
02 Std. Sol.	0.019	16:34
03 Room Air	0.000	16:34
04 Std. Sol.	0.020	16:35
05 Room Air	0.000	16:36
06 Std. Sol.	0.020	16:36
07 Room Air	0.000	16:37

08 Sim Temp = 34.0°C

Simul Ser No = MP3064
Std Sol No = 20070
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.14.00 09/16
04/13/2021 16:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:43
02 Std. Sol.	0.304	16:44
03 Room Air	0.000	16:44
04 Std. Sol.	0.305	16:45
05 Room Air	0.000	16:46
06 Std. Sol.	0.305	16:46
07 Room Air	0.000	16:47

08 Sim Temp = 34.0°C

Simul Ser No = MP5317
Std Sol No = 202012A
County = 08 Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC

0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.14.00 09/16
04/13/2021 16:49

Test	AC	Time
01 Room Air	0.000	16:50
02 *Subject Test	INT*	16:51
03 Room Air	0.000	16:51

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 6
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AE + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.14.00 09/16
04/13/2021 16:51

Test	AC	Time
01 Diagnostic	OK	16:52
02 Room Air	0.000	16:53
03 *Subject Test	RFI*	16:53
04 Room Air	0.000	16:54

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 6
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.14.00 09/16
04/13/2021 16:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:56
02 Std. Gas	0.082	16:57
03 Room Air	0.000	16:57
04 Std. Gas	0.083	16:57
05 Room Air	0.000	16:58
06 Std. Gas	0.083	16:58
07 Room Air	0.000	16:59

Lot No = 24119080A1
Cyl No = 6
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.14.00 09/16
04/13/2021 16:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:00
02 Std. Gas	0.082	17:00
03 Room Air	0.000	17:00
04 Std. Gas	0.082	17:01
05 Room Air	0.000	17:01
06 Std. Gas	0.082	17:02
07 Room Air	0.000	17:02

Lot No = 24119080A1
Cyl No = 6
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.14.00 09/16
04/13/2021 17:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:03
02 Std. Gas	0.082	17:03
03 Room Air	0.000	17:04
04 Std. Gas	0.083	17:04
05 Room Air	0.000	17:04
06 Std. Gas	0.083	17:05
07 Room Air	0.000	17:05

Lot No = 24119080A1
Cyl No = 6
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Form 106-I8000