

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006496

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 830 psi Regulator: 825 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
  - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.015 AC  
Sim. Ser #: MP5319 Lot #: 201911C Exp. Date: 11-19-21
  - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.400 AC  
Sim. Ser #: MP3062 Lot #: 202103E Exp. Date: 3-24-23

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10AC + 0.05% Acetone  
Sim. Ser #: DR5114 Lot #: ICS 6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 2411908041 Cyl #: 5 Exp. Date: 11-5-21

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.080</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED EXTERNAL BATTERY, TAPED PROTECTIVE SLEEVE ON BREATH HOSE DUE TO TEAR

Instrument is acceptable to be used in the field.

Charles E. Eden  
Breath Analyst Signature

4.14.21  
Date

Kalish H. Hiest  
Reviewed by

4-19-21  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006496  
Location = TOXL      8164.13.00 06/09  
04/14/2021      11:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:36
02 Std. Sol.	0.014	11:37
03 Room Air	0.000	11:37
04 Std. Sol.	0.014	11:38
05 Room Air	0.000	11:39
06 Std. Sol.	0.014	11:39
07 Room Air	0.000	11:40

08 Sim Temp = 34.0°C

Simul Ser No = MP5319  
Std Sol No = 201911C  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Low AC*  
*0.015 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006496  
Location = TOXL                      8164.13.00 06/09  
04/14/2021                              11:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:47
02 Std. Sol.	0.402	11:48
03 Room Air	0.000	11:48
04 Std. Sol.	0.401	11:49
05 Room Air	0.000	11:50
06 Std. Sol.	0.402	11:51
07 Room Air	0.000	11:51

08 Sim Temp = 34.0°C

Simul Ser No = MP3062  
Std Sol No = 202103E  
County = 08                              Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:      HIGH AC  
                    0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006496  
Location = TOXL                    8164.13.00 06/09  
04/14/2021                            11:52

Test	AC	Time
01 Room Air	0.000	11:54
02 *Subject Test	INT*	11:54
03 Room Air	0.000	11:55

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/08/1987  
Sub Sex = Male                    Weight = 150  
Test = DUI                        Cit = INTERFERENT  
Dr. Lic. = ND/DIS871456  
Lot No = 24119080A1  
Cyl No = 5  
Expiration Date = 11/05/2021  
County = 08                        Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006496  
Location = TOXL      8164.13.00 06/09  
04/14/2021      11:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:59
02 Std. Gas	0.081	11:59
03 Room Air	0.000	12:00
04 Std. Gas	0.081	12:00
05 Room Air	0.000	12:01
06 Std. Gas	0.081	12:01
07 Room Air	0.000	12:02

Lot No = 24119080A1  
Cyl No = 5  
Exp Date = 11/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006496  
Location = TOXL      8164.13.00 06/09  
04/14/2021      12:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:03
02 Std. Gas	0.081	12:03
03 Room Air	0.000	12:04
04 Std. Gas	0.081	12:04
05 Room Air	0.000	12:04
06 Std. Gas	0.081	12:05
07 Room Air	0.000	12:05

Lot No = 24119080A1  
Cyl No = 5  
Exp Date = 11/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080AC

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006496  
Location = TOXL      8164.13.00 06/09  
04/14/2021      12:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:06
02 Std. Gas	0.081	12:07
03 Room Air	0.000	12:07
04 Std. Gas	0.081	12:07
05 Room Air	0.000	12:08
06 Std. Gas	0.080	12:08
07 Room Air	0.000	12:09

Lot No = 24119080A1  
Cyl No = 5  
Exp Date = 11/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000