

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004938

Location: TOXL

A. General Setup and Checks:

- 1. Diagnostic Tests Pass and Instrument in Ready Mode
- 2. Breath tube heated
- 3. Date, time and location code (Level 2,E). Re-set if necessary.
- 4. Print test (Level 1,P). Sign and attach test record.
- 5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 148 psi Regulator: 125 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3064 Lot #: 20070 Exp. Date: 2.13.22
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP5317 Lot #: 202012A Exp. Date: 12.8.22

Configure dry gas standard for the following tests (Level 1,S).

- 2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: ICS6 Exp. Date: NA
 Display reads "Interferent Detect"
- 3. RFI Check (CMS Mode) Display reads "RFI Detect"
- 4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 24119080A1 Cyl #: 6 Exp. Date: 11.5.21

| | | | | | |
|---------|--------------|--------|--------------|--------|--------------|
| Test 1 | <u>0.079</u> | Test 4 | <u>0.081</u> | Test 7 | <u>0.080</u> |
| Test 2 | <u>0.079</u> | Test 5 | <u>0.081</u> | Test 8 | <u>0.080</u> |
| Test 3 | <u>0.080</u> | Test 6 | <u>0.080</u> | Test 9 | <u>0.079</u> |
| Average | <u>0.080</u> | | | | |

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eck
Breath Analyst Signature

4.15.21
Date

Kali L. Hiett
Reviewed by

4-19-21
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/15/2021 09:00

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 6
Standard Expiration? 11/05/2021
Oper No? 666666

Flow Cal. Date: 06/10/2020
Slope 587
Intercept -338327

IR Calibration Date: 06/10/2020
 3um 9um

0th Coef(*100): -22434 -23179
1st Coef(*100): 269849 136790
2nd Coef(*100): 963 840
H2O adj(mg/l*10k): 355 370

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/15/2021 09:02

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 09:03 |
| 02 Std. Sol. | 0.019 | 09:04 |
| 03 Room Air | 0.000 | 09:04 |
| 04 Std. Sol. | 0.020 | 09:05 |
| 05 Room Air | 0.000 | 09:05 |
| 06 Std. Sol. | 0.020 | 09:06 |
| 07 Room Air | 0.000 | 09:07 |

08 Sim Temp = 34.0°C

Simul Ser No = MP3064
Std Sol No = 20070
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/15/2021 09:38

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 09:38 |
| 02 Std. Sol. | 0.302 | 09:39 |
| 03 Room Air | 0.000 | 09:40 |
| 04 Std. Sol. | 0.302 | 09:40 |
| 05 Room Air | 0.000 | 09:41 |
| 06 Std. Sol. | 0.300 | 09:42 |
| 07 Room Air | 0.000 | 09:42 |

08 Sim Temp = 34.0°C

Simul Ser No = MP5317
Std Sol No = 202012A
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/15/2021 10:10

| Test | AC | Time |
|------------------|-------|-------|
| 01 Diagnostic | OK | 10:11 |
| 02 Room Air | 0.000 | 10:11 |
| 03 *Subject Test | RFI* | 10:11 |
| 04 Room Air | 0.000 | 10:12 |

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 6
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: RFI CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/15/2021 10:12

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:13 |
| 02 Std. Gas | 0.079 | 10:13 |
| 03 Room Air | 0.000 | 10:13 |
| 04 Std. Gas | 0.079 | 10:14 |
| 05 Room Air | 0.000 | 10:14 |
| 06 Std. Gas | 0.080 | 10:15 |
| 07 Room Air | 0.000 | 10:15 |

Lot No = 24119080A1
Cyl No = 6
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/15/2021 10:17

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:18 |
| 02 Std. Gas | 0.081 | 10:18 |
| 03 Room Air | 0.000 | 10:19 |
| 04 Std. Gas | 0.081 | 10:19 |
| 05 Room Air | 0.000 | 10:19 |
| 06 Std. Gas | 0.080 | 10:20 |
| 07 Room Air | 0.000 | 10:20 |

Lot No = 24119080A1
Cyl No = 6
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

