

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005952

Location: TOXL

A. General Setup and Checks:

- 1.  Diagnostic Tests Pass and Instrument in Ready Mode
- 2.  Breath tube heated
- 3.  Date, time and location code (Level 2,E). Re-set if necessary.
- 4.  Print test (Level 1,P). Sign and attach test record.
- 5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display 253 psi Regulator: 250 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.030 AC  
Sim. Ser #: MP5289 Lot #: 201911 E Exp. Date: 11.19.21
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.400 AC  
Sim. Ser #: MP3062 Lot #: 202103 E Exp. Date: 3.24.23

Configure dry gas standard for the following tests (Level 1,S).

- 2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3847 Lot #: ICS 6 Exp. Date: NA  
 Display reads "Interferent Detect"
- 3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
- 4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 02621080A | Cyl #: 25 Exp. Date: 3.5.23  
Test 1 0.080 Test 4 0.080 Test 7 0.080  
Test 2 0.080 Test 5 0.080 Test 8 0.080  
Test 3 0.080 Test 6 0.081 Test 9 0.081  
Average 0.080

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

[Signature]  
Breath Analyst Signature

4.19.21  
Date

[Signature]  
Reviewed by

4-19-21  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005952  
Location = TOXL      8164.14.00 09/16  
04/19/2021      11:51

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#%&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#%&\*()\_+?

Current Instrument Setup


Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      02621080A1  
Standard Cyl #?      25  
Standard Expiration?      03/05/2023  
Oper No?      666666

Flow Cal. Date:      04/16/2021  
Slope      657  
Intercept      -744615

IR Calibration Date:      04/16/2021

	3um	9um
0th Coef(*100):	-24450	-14639
1st Coef(*100):	266154	132500
2nd Coef(*100):	2284	1091
H2O adj(mg/l*10k):	516	323

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature

CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005952  
Location = TOXL      8164.14.00 09/16  
04/19/2021      11:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:56
02 Std. Sol.	0.028	11:57
03 Room Air	0.000	11:57
04 Std. Sol.	0.028	11:58
05 Room Air	0.000	11:59
06 Std. Sol.	0.028	11:59
07 Room Air	0.000	12:00

08 Sim Temp = 34.0°C

Simul Ser No = MP5289  
Std Sol No = 201911E  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.030 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005952  
Location = TOXL      8164.14.00 09/16  
04/19/2021      12:02

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:02
02 Std. Sol.	0.394	12:03
03 Room Air	0.000	12:04
04 Std. Sol.	0.395	12:04
05 Room Air	0.000	12:05
06 Std. Sol.	0.395	12:06
07 Room Air	0.000	12:06

08 Sim Temp = 34.0°C

Simul Ser No = MP3062  
Std Sol No = 202103E  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-005952  
Location = TOXL                              8164.14.00 09/16  
04/19/2021                                              12:08

Test	AC	Time
01 Room Air	0.000	12:09
02 *Subject Test	INT*	12:10
03 Room Air	0.000	12:10

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                              Weight = 150  
Test = DUI                                      Cit = INTERFERENT  
Dr. Lic. = ND/DIS921456  
Lot No = 02621080A1  
Cyl No = 25  
Expiration Date = 03/05/2023  
County = 08                                      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                    8164.14.00 09/16  
04/19/2021                            12:11

Test	AC	Time
01 Diagnostic	OK	12:11
02 Room Air	0.000	12:12
03 *Subject Test	RFI*	12:12
04 Room Air	0.000	12:12

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                      Weight = 150  
Test = DUI                              Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 02621080A1  
Cyl No = 25  
Expiration Date = 03/05/2023  
County = 08                              Oper No. = 666666

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                 8164.14.00 09/16  
04/19/2021                        12:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:14
02 Std. Gas	0.080	12:14
03 Room Air	0.000	12:15
04 Std. Gas	0.080	12:15
05 Room Air	0.000	12:16
06 Std. Gas	0.080	12:16
07 Room Air	0.000	12:17

Lot No = 02621080A1

Cyl No = 25

Exp Date = 03/05/2023

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005952  
Location = TOXL      8164.14.00 09/16  
04/19/2021      12:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:17
02 Std. Gas	0.080	12:18
03 Room Air	0.000	12:18
04 Std. Gas	0.080	12:19
05 Room Air	0.000	12:19
06 Std. Gas	0.081	12:19
07 Room Air	0.000	12:20

Lot No = 02621080A1  
Cyl No = 25  
Exp Date = 03/05/2023  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005952  
Location = TOXL      8164.14.00 09/16  
04/19/2021      12:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:21
02 Std. Gas	0.080	12:21
03 Room Air	0.000	12:22
04 Std. Gas	0.080	12:22
05 Room Air	0.000	12:23
06 Std. Gas	0.081	12:23
07 Room Air	0.000	12:23

Lot No = 02621080A1  
Cyl No = 25  
Exp Date = 03/05/2023  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-18000