

INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005952

A. GENERAL Setup and Checks:

- Power On
 - 5 min. countdown during initial warm-up
 - 1 min. countdown to begin test after standby mode
- Display Intensity (Ctrl + ↑ or ↓)
3. Password Security:
 - Level 1 (Operator) Level 2 (Field Inspector) Level 3 (Lab Personnel)
- Printer Setup (Level 3,S,P)
- Forms Load (CMI, Inc., CoBrA Software)

B. LEVEL 1 Setup and Checks:

- Internal Printer Test and Paper Feed (Level 1,P). Attach record.
- External Printer Test (Level 1,P). Attach record.
3. Barcode Reader (Level 1,B):
 - Operator ID ND License R key: Refusal. Attach record.

C. LEVEL 2 Setup and Checks:

- Continuous Air Blank (Level 2,A)
- Diagnostic (Level 2,D). Attach record.
- Prelim Data Entry (Level 2,E)
4. Ser Number/Version Info (Level 2,V):
 - Inst Ser#: 80-005952 DSP Ver: 1362.14
 - Core Ver: 8164.14 CPLD Ver: 7

D. LEVEL 3 Setup and Checks:

- DVM Monitor (Level 3,D,D): 3: 12729 s: 0 c: 47.06
9: 12979 s: 7 B: 45 0
(RFI indicator with keyed radio): ✓ 1
2. Function (F) Keys: : F5 – Tone F9 - Fan
 F6 – Green LED F10 – Pump
 F7 – Red LED F11 – Solenoid
3. Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 963 CF: 1.05
4. Flow Monitor (Level 3,D,F): Flow Sensor, R: 218 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.0 I: 0.0
B: 45.1 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 237 psi P (gauge): 250 psi
7. General Setup (Level 3,S,G)
8. Configure Start Test (Level 3,S,T)
9. Network (Level 3,S,N)

E. TESTS (attach all test records):

Configure wet bath simulator standards before running tests (Level 1,S):

1. Zero AC (Level 1,C):
Sim ser #: MP5320 Lot #: MILLI-Q H₂O Exp date: NA
Known value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):
Sim ser #: MP5319 Lot #: 201911C Exp date: 11.19.21
Known value: 0.015 AC
Results: 0.015 AC 0.015 AC 0.015 AC
3. High AC (Level 1,C):
Sim ser #: MP3061 Lot #: 20101 Exp date: 2.13.22
Known value: 0.500 AC
Results: 0.496 AC 0.497 AC 0.499 AC

Configure dry gas standard before running tests (Level 1,S):

4. Ethanol + 0.05% Acetone (Level 1,B):
Sim ser #: DR5114 Lot #: ICS 6 Exp date: NA
Known value: 0.10 AC + 0.05% ACETONE
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):
Sim ser #: DR5188 Lot #: ACE 1 Exp date: NA
Known value: 0.05% ACETONE
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):
Sim ser #: DR5106 Lot #: MET 3 Exp date: NA
Known value: 0.10% METHANOL
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):
Sim ser #: DR5189 Lot #: IPA 2 Exp date: NA
Known value: 0.10% ISOPROPANOL
 Display reads "Interferent Detect"

- 8. Ambient Fail (Level 1,B): Display and test record read correct.
- 9. Invalid Sample (Level 1,B): Display and test record read correct.
- 10. Sequence Aborted (Level 1,B): Display and test record read correct.
- 11. ABA (Level 1,B): Test record reads correct.
- 12. Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13. RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14. Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15. Deficient Samples (CMS mode): Display and test record read correct.
- 16. Improper Sample (CMS mode): Display and test record read correct.
- 17. Difference Too Great (CMS mode): Display and test record read correct.
- 18. CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):


Ave: 0.0805 Std Dev: 0.0007 Rel Std Dev: 0.86%

Maintenance/Repairs/Notes: INITIAL INSPECTION DUE TO
REPLACING ANALYTICAL CHAMBER ASSEMBLY
FOR A REPAIR.

Instrument is acceptable to be used in the field.


Field Inspector's Signature

4-19-21
Date


Reviewed by

4-19-21
Date

Intoxilyzer Test Record and Checklist
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 09:03

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMN

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMN

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMN

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMN


Current Instrument Setup
Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On MemFull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 6
Standard Expiration? 11/05/2021
Oper No? 666666

Flow Cal. Date: 04/16/2021
Slope 657
Intercept -744615

IR Calibration Date: 04/16/2021

	3um	9um
0th Coef(*100):	-24450	-14639
1st Coef(*100):	266154	132500
2nd Coef(*100):	2284	1091
H2O adj(mg/l*10K):	516	323

***** Test End *****



Operator Signature
CHARLES EDER

Remarks:

*Internal
PRINT
TEST*

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 09:04

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 6
Standard Expiration? 11/05/2021
Oper No? 666666

Flow Cal. Date: 04/16/2021
Slope 657
Intercept -744615

IR Calibration Date: 04/16/2021
 3um 9um

0th Coef(*100): -24450 -14639
1st Coef(*100): 266154 132500
2nd Coef(*100): 2284 1091
H2O adj(mg/l*10k): 516 323

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

EXTERNAL
PRINT TEST

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 09:06

Test	AC	Time
01 Room Air	0.000	09:06
02 *Subject Test	REF*	09:07
03 Room Air	0.000	09:08

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = R KEY REFUSAL
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 6
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: "R" KEY REFUSAL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 09:13

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *DIAGNOSTIC*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 09:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:21
02 Std. Sol.	0.000	09:22
03 Room Air	0.000	09:23
04 Std. Sol.	0.000	09:23
05 Room Air	0.000	09:24
06 Std. Sol.	0.000	09:24
07 Room Air	0.000	09:25

08 Sim Temp = 34.0°C

Simul Ser No = MP5320
Std Sol No = MILLIQ H2O
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: ZERO AC
 0.000 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 09:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:28
02 Std. Sol.	0.015	09:29
03 Room Air	0.000	09:29
04 Std. Sol.	0.015	09:30
05 Room Air	0.000	09:30
06 Std. Sol.	0.015	09:31
07 Room Air	0.000	09:32


08 Sim Temp = 34.0°C

Simul Ser No = MP5319

Std Sol No = 201911C

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC
0.015 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 09:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:34
02 Std. Sol.	0.496	09:35
03 Room Air	0.000	09:35
04 Std. Sol.	0.497	09:36
05 Room Air	0.000	09:37
06 Std. Sol.	0.499	09:37
07 Room Air	0.000	09:38


08 Sim Temp = 34.0°C

Simul Ser No = MP3061

Std Sol No = 20101

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.500 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 09:41

Test	AC	Time
01 Room Air	0.000	09:41
02 *Subject Test	INT*	09:42
03 Room Air	0.000	09:43

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 09:45

Test	AC	Time
01 Room Air	0.000	09:48
02 *Subject Test	INT*	09:49
03 Room Air	0.000	09:50

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.05% ACETONE

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 10:00

Test	AC	Time
01 Room Air	0.000	10:01
02 *Subject Test	INT*	10:01
03 Room Air	0.000	10:02

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10% METHANOL

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 10:02

Test	AC	Time
01 Room Air	0.000	10:03
02 *Subject Test	INT*	10:03
03 Room Air	0.000	10:04

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10% ISOPROPAOL

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 10:07

Test	AC	Time
01 Room Air	AMB*	10:08
02 Room Air	0.000	10:09

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = AMBIENT FAIL
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *AMBIENT FAIL*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 10:17

Test	AC	Time
01 Room Air	0.000	10:18
02 Invalid Sample	X.XXX	10:18
03 Room Air	0.000	10:19

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INVALID SAMPLE
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INVALID SAMPLE
MOUTH ALCOHOL

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 10:20

Test	AC	Time
01 Room Air	0.000	10:21
02 *Subject Test	ABT*	10:21
03 Room Air	0.000	10:21

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = SEQ ABORTED
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: SEQUENCE ABORTED

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 10:21

Test	AC	Time
01 Room Air	0.000	10:23
02 Subject Test 1	0.150	10:23
03 Room Air	0.000	10:24
04 Reported AC	0.150	10:23

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = ABA MODE
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ABA MODE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 10:25

Test	AC	Time
01 Diagnostic	OK	10:26
02 Room Air	0.000	10:27
03 Subject Test 1	0.149	10:27
04 Room Air	0.000	10:29
05 Std. Gas	0.088*	10:30
06 Room Air	0.000	10:31

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = CAL CHK OUT TOL
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
OUT OF TOLERANCE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 10:34

Test	AC	Time
01 Diagnostic	OK	10:35
02 Room Air	0.000	10:36
03 *Subject Test	RFI*	10:36
04 Room Air	0.000	10:36

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 10:37

Test	AC	Time
01 Diagnostic	OK	10:38
02 Room Air	0.000	10:38
03 *Subject Test	0.000*	10:41
04 Room Air	0.000	10:44
05 Std. Gas	0.080	10:45
06 Room Air	0.000	10:46
07 *Subject Test	0.000*	10:49
08 Room Air	0.000	10:50

No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = DEFICIENT SAMP
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

DEFICIENT
SAMPLES

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 10:54

Test	AC	Time
01 Diagnostic	OK	10:55
02 Room Air	0.000	10:55
03 Subject Test 1	0.000	10:55
04 Room Air	IPS*	10:56
05 Room Air	0.000	10:57

*Improper Sample

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = IMPROPER SAMPLE
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

IMPROPER SAMPLE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 10:57

Test	AC	Time
01 Diagnostic	OK	10:58
02 Room Air	0.000	10:58
03 Subject Test 1	0.197*	10:59
04 Room Air	0.000	11:01
05 Std. Gas	0.080	11:02
06 Room Air	0.000	11:03
07 Subject Test 2	0.149*	11:04
08 Room Air	0.000	11:05

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = DIFF TOO GREAT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: DIFFERENCE
 TOO GREAT

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 11:06


Test	AC	Time
01 Diagnostic	OK	11:07
02 Room Air	0.000	11:07
03 Subject Test 1	0.150	11:07
04 Room Air	0.000	11:10
05 Std. Gas	0.079	11:11
06 Room Air	0.000	11:12
07 Subject Test 2	0.150	11:13
08 Room Air	0.000	11:14
09 Reported AC	0.150	11:13

Difference OK
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = CMS MODE
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 48
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *CMS MODE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = TOXL 8164.14.00 09/16
04/19/2021 11:14

Test	AC	Time
01 Room Air	0.000	11:15
02 Std. Gas	0.079	11:15
03 Room Air	0.000	11:16
04 Std. Gas	0.080	11:16
05 Room Air	0.000	11:17
06 Std. Gas	0.080	11:17
07 Room Air	0.000	11:18
08 Std. Gas	0.080	11:18
09 Room Air	0.000	11:18
10 Std. Gas	0.080	11:19
11 Room Air	0.000	11:19
12 Std. Gas	0.079	11:20
13 Room Air	0.000	11:20
14 Std. Gas	0.080	11:21
15 Room Air	0.000	11:21
16 Std. Gas	0.081	11:21
17 Room Air	0.000	11:22
18 Std. Gas	0.081	11:22
19 Room Air	0.000	11:23
20 Std. Gas	0.081	11:23
21 Room Air	0.000	11:24
22 Std. Gas	0.080	11:24
23 Room Air	0.000	11:25
24 Std. Gas	0.081	11:25
25 Room Air	0.000	11:25
26 Std. Gas	0.081	11:26
27 Room Air	0.000	11:26
28 Std. Gas	0.081	11:27
29 Room Air	0.000	11:27
30 Std. Gas	0.081	11:28
31 Room Air	0.000	11:28
32 Std. Gas	0.081	11:28
33 Room Air	0.000	11:29
34 Std. Gas	0.081	11:29
35 Room Air	0.000	11:30
36 Std. Gas	0.081	11:30
37 Room Air	0.000	11:31
38 Std. Gas	0.080	11:31
39 Room Air	0.000	11:32
40 Std. Gas	0.081	11:32
41 Room Air	0.000	11:33
42 Cal Check Stats:		
43 Average	0.0805	
44 Std Dev	0.0007	
45 Rel Std Dev	0.86%	



Operator Signature
CHARLES EDER

Remarks: 0.080 AC
Lot: 05620080 A1
Exp: 4.5.22

Form 106-I8000

C/L: 048