

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005945

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 641 psi Regulator: 625 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
  - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP3064 Lot #: 20070 Exp. Date: 2.13.22
  - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MPS317 Lot #: 202012A Exp. Date: 12.8.22

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3847 Lot #: FCS6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 02621080A1 Cyl #: 25 Exp. Date: 3.5.23

Test 1 <u>0.079</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.080</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>
Average <u>0.079</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION

---

---

---

Instrument is acceptable to be used in the field.

Charles E. Ed  
Breath Analyst Signature

4.28.21  
Date

Karl L. Hib  
Reviewed by

5-3-21  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005945  
Location = TOXL                8164.14.00 09/16  
04/28/2021                        16:19

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            02621080A1  
Standard Cyl #?            25  
Standard Expiration?       03/05/2023  
Oper No?                    666666

Flow Cal. Date:            06/20/2013  
Slope                        637  
Intercept                   -566407

IR Calibration Date:       05/22/2015

	3um	9um
0th Coef(*100):	-21213	-23093
1st Coef(*100):	279171	139292
2nd Coef(*100):	2898	1291
H2O adj(mg/l*10k):	332	319

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/28/2021      16:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:23
02 Std. Sol.	0.019	16:24
03 Room Air	0.000	16:25
04 Std. Sol.	0.019	16:25
05 Room Air	0.000	16:26
06 Std. Sol.	0.018	16:27
07 Room Air	0.000	16:27

08 Sim Temp = 34.0°C

Simul Ser No = MP3064  
Std Sol No = 20070  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-005945  
Location = TOXL                                  8164.14.00 09/16  
04/28/2021    16:33

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:34
02 Std. Sol.	0.304	16:35
03 Room Air	0.000	16:35
04 Std. Sol.	0.304	16:36
05 Room Air	0.000	16:37
06 Std. Sol.	0.303	16:37
07 Room Air	0.000	16:38

08 Sim Temp = 34.0°C

Simul Ser No = MP5317  
Std Sol No = 202012A  
County = 08    Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*HIGH AC*  
*0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/28/2021      16:39

Test	AC	Time
01 Room Air	0.000	16:40
02 *Subject Test	INT*	16:40
03 Room Air	0.000	16:41

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS921456  
Lot No = 02621080A1  
Cyl No = 25  
Expiration Date = 03/05/2023  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer                  Alcohol Analyzer  
North Dakota Model 8000              SN 80-005945  
Location = TOXL                        8164.14.00 09/16  
04/28/2021                                16:41

Test	AC	Time
01 Diagnostic	OK	16:42
02 Room Air	0.000	16:43
03 *Subject Test	RFI*	16:43
04 Room Air	0.000	16:43

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                        Weight = 150  
Test = DUI                                Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 02621080A1  
Cyl No = 25  
Expiration Date = 03/05/2023  
County = 08                                Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/28/2021      16:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:44
02 Std. Gas	0.079	16:45
03 Room Air	0.000	16:45
04 Std. Gas	0.079	16:45
05 Room Air	0.000	16:46
06 Std. Gas	0.079	16:46
07 Room Air	0.000	16:47

Lot No = 02621080A1  
Cyl No = 25  
Exp Date = 03/05/2023  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/28/2021      16:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:48
02 Std. Gas	0.080	16:48
03 Room Air	0.000	16:49
04 Std. Gas	0.079	16:49
05 Room Air	0.000	16:50
06 Std. Gas	0.079	16:50
07 Room Air	0.000	16:50

Lot No = 02621080A1  
Cyl No = 25  
Exp Date = 03/05/2023  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/28/2021      16:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:51
02 Std. Gas	0.080	16:52
03 Room Air	0.000	16:52
04 Std. Gas	0.080	16:52
05 Room Air	0.000	16:53
06 Std. Gas	0.080	16:53
07 Room Air	0.000	16:54

Lot No = 02621080A1  
Cyl No = 25  
Exp Date = 03/05/2023  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000