

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005358

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 317 psi Regulator: 300 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3064 Lot #: 20070 Exp. Date: 2.13.22
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP5317 Lot #: 128 Exp. Date: 12.8.22
CEE 202012A

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: ICS 6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 25 Exp. Date: 3.5.23

Test 1	<u>0.081</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.082</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.080</u>	Test 6	<u>0.081</u>	Test 9	<u>0.081</u>
Average	<u>0.081</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED EXTERNAL BATTERY,

Instrument is acceptable to be used in the field.

Breath Analyst Signature

Reviewed by

Date

Date

5-4-21

5-4-21

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005358
Location = TOXL 8164.14.00 09/16
05/04/2021 11:54

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 25
Standard Expiration? 03/05/2023
Oper No? 666666

Flow Cal. Date: 06/04/2019
Slope 570
Intercept 4079

IR Calibration Date: 06/04/2019

	3um	9um
0th Coef (*100):	-19471	-24652
1st Coef (*100):	273237	143582
2nd Coef (*100):	2965	1293
H2O adj (mg/l*10k):	273	374

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005358
Location = TOXL 8164.14.00 09/16
05/04/2021 11:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:57
02 Std. Sol.	0.019	11:58
03 Room Air	0.000	11:58
04 Std. Sol.	0.018	11:59
05 Room Air	0.000	12:00
06 Std. Sol.	0.019	12:00
07 Room Air	0.000	12:01

08 Sim Temp = 34.0°C

Simul Ser No = MP3064
Std Sol No = 20070
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005358
Location = TOXL 8164.14.00 09/16
05/04/2021 12:02

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:03
02 Std. Sol.	0.302	12:03
03 Room Air	0.000	12:04
04 Std. Sol.	0.303	12:05
05 Room Air	0.000	12:05
06 Std. Sol.	0.302	12:06
07 Room Air	0.000	12:07

08 Sim Temp = 34.0°C

Simul Ser No = MP5317
Std Sol No = 202012A
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005358
Location = TOXL 8164.14.00 09/16
05/04/2021 12:09

Test	AC	Time
01 Room Air	0.000	12:10
02 *Subject Test	INT*	12:10
03 Room Air	0.000	12:11

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/08/1987
Sub Sex = Male Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS871456
Lot No = 02621080A1
Cyl No = 25
Expiration Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005358
Location = TOXL 8164.14.00 09/16
05/04/2021 12:11

Test	AC	Time
01 Diagnostic	OK	12:12
02 Room Air	0.000	12:12
03 *Subject Test	RFI*	12:13
04 Room Air	0.000	12:13

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/08/1987
Sub Sex = Male Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS871456
Lot No = 02621080A1
Cyl No = 25
Expiration Date = 03/05/2023
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005358
Location = TOXL 8164.14.00 09/16
05/04/2021 12:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:14
02 Std. Gas	0.081	12:14
03 Room Air	0.000	12:15
04 Std. Gas	0.082	12:15
05 Room Air	0.000	12:16
06 Std. Gas	0.080	12:16
07 Room Air	0.000	12:16

Lot No = 02621080A1
Cyl No = 25
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005358
Location = TOXL 8164.14.00 09/16
05/04/2021 12:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:18
02 Std. Gas	0.081	12:18
03 Room Air	0.000	12:19
04 Std. Gas	0.081	12:19
05 Room Air	0.000	12:19
06 Std. Gas	0.081	12:20
07 Room Air	0.000	12:20

Lot No = 02621080A1
Cyl No = 25
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005358
Location = TOXL 8164.14.00 09/16
05/04/2021 12:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:21
02 Std. Gas	0.081	12:22
03 Room Air	0.000	12:22
04 Std. Gas	0.081	12:23
05 Room Air	0.000	12:23
06 Std. Gas	0.081	12:24
07 Room Air	0.000	12:24

Lot No = 02621080A1
Cyl No = 25
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC

Form 106-I8000