



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-005947	Instrument Location Dickinson LEC
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Annual inspection performed at crime laboratory by Charles Eder</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): NONE
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Randy Brown</i> 4655	Date 05/06/2021
--	--------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Charles Eder</i>	Certified Date 5.17.21
--	---------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
05/06/2021 20:18

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 32
Standard Expiration? 03/05/2023
Oper No? 132406

Flow Cal. Date: 06/21/2013
Slope 693
Intercept -735360

IR Calibration Date: 04/06/2015
 3um 9um

0th Coef(*100): -28096 -26982
1st Coef(*100): 280034 136042
2nd Coef(*100): 3530 1567
H2O adj(mg/l*10k): 510 437

***** Printer Test End *****

 4835

Operator Signature
RUSSELL PRIEWE

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
05/06/2021 20:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:23
02 Std. Gas	0.080	20:23
03 Room Air	0.000	20:24
04 Std. Gas	0.079	20:24
05 Room Air	0.000	20:25
06 Std. Gas	0.079	20:25
07 Room Air	0.000	20:26

Lot No = 02621080A1
Cyl No = 32
Exp Date = 03/05/2023
County = 45 Oper No. = 132406


Operator Signature
RUSSELL PRIEWE

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
05/06/2021 20:27

Test	AC	Time
01 Room Air	0.000	20:29
02 *Subject Test	RFI*	20:30
03 Room Air	0.000	20:30

*Invalid Test
Inhibited - RFI

Russell Priewe 4/55

Operator Signature
RUSSELL PRIEWE

Remarks:

Form 106-I8000

NOTE: OFFICER PERFORMED RFI TEST IN
ABA MODE. NO SUBJECT OR GAS
CYLINDER INFORMATION APPEAR BECAUSE
OFFICER DID NOT ENTER FIRST NAME
OR LAST NAME FOR SUBJECT.
RFI TEST IS STILL GOOD.

Charles E. Eh
5.17.21