



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <i>80-004940</i>	Instrument Location <i>N.S.</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <i>5/11/2021</i>
-------------------------------	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date <i>5/24/2021</i>
---------------------------	------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004940  
Location = NS                    8164.14.00 09/16  
05/11/2021                        10:11

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            24119080A1  
Standard Cyl #?            16  
Standard Expiration?       11/05/2021  
Oper No?                    020611

Flow Cal. Date:            04/06/2020  
Slope                        635  
Intercept                   -585462

IR Calibration Date:        04/06/2020

	3um	9um
0th Coef(*100):	-17993	-22378
1st Coef(*100):	260246	135868
2nd Coef(*100):	2982	1272
H2O adj(mg/l*10k):	383	361

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
TOM SHARP

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-004940  
Location = NS                                      8164.14.00 09/16  
05/11/2021    10:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:13
02 Std. Gas	0.080	10:13
03 Room Air	0.000	10:14
04 Std. Gas	0.079	10:14
05 Room Air	0.000	10:15
06 Std. Gas	0.079	10:15
07 Room Air	0.000	10:15

Lot No = 24119080A1  
Cyl No = 16  
Exp Date = 11/05/2021  
County = 30                                      Oper No. = 020611

  
Operator Signature  
TOM SHARP

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004940  
Location = NS      8164.14.00 09/16  
05/11/2021      10:07

Test	AC	Time
01 Room Air	0.000	10:09
02 *Subject Test	RFI*	10:09
03 Room Air	0.000	10:09

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, TEST TEST  
Sub DOB = 01/01/2000  
Sub Sex = Male      Weight = 150  
Test = DUI      Cit = 000000  
Dr. Lic. = ND/000000  
Lot No = 24119080A1  
Cyl No = 16  
Expiration Date = 11/05/2021  
County = 30      Oper No. = 020611

  
Operator Signature  
TOM SHARP

Remarks:

Form 106-I8000