

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006494 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 999 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.010 AC
Sim. Ser #: MP3070 Lot #: 201906A Exp. Date: 6-11-21
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.500 AC
Sim. Ser #: MP3061 Lot #: 20101 Exp. Date: 2-13-22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: ICS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 22 Exp. Date: 3-5-23

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.080</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED
SIMULATOR RETURN O-RING

Instrument is acceptable to be used in the field.

[Signature]
Breath Analyst Signature

5/20/21
Date

[Signature]
Reviewed by

5/24/21
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.14.00 09/16
05/20/2021 10:58

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 22
Standard Expiration? 03/05/2023
Oper No? 666666

Flow Cal. Date: 08/18/2015
Slope 655
Intercept -624238

IR Calibration Date: 05/10/2017

	3um	9um
0th Coef(*100):	-28969	-21478
1st Coef(*100):	256589	140860
2nd Coef(*100):	2527	1330
H2O adj(mg/l*10k):	655	473

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: **PRINT TEST**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.14.00 09/16
05/20/2021 11:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:10
02 Std. Sol.	0.008	11:10
03 Room Air	0.000	11:11
04 Std. Sol.	0.009	11:12
05 Room Air	0.000	11:12
06 Std. Sol.	0.008	11:13
07 Room Air	0.000	11:13


08 Sim Temp = 34.0°C

Simul Ser No = MP3070

Std Sol No = 201906A

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC
0.010 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.14.00 09/16
05/20/2021 11:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:17
02 Std. Sol.	0.497	11:18
03 Room Air	0.000	11:18
04 Std. Sol.	0.496	11:19
05 Room Air	0.000	11:20
06 Std. Sol.	0.496	11:21
07 Room Air	0.000	11:21

08 Sim Temp = 34.0°C

Simul Ser No = MP3061

Std Sol No = 20101

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.500 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.14.00 09/16
05/20/2021 11:25

Test	AC	Time
01 Room Air	0.000	11:26
02 *Subject Test	INT*	11:26
03 Room Air	0.000	11:27

*Invalid Test
Interferent Detected

Sub Name = TEST, BOBBY NONE
Sub DOB = 01/01/1985
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/TES851012
Lot No = 02621080A1
Cyl No = 22
Expiration Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.14.00 09/16
05/20/2021 11:27

Test	AC	Time
01 Diagnostic	OK	11:28
02 Room Air	0.000	11:28
03 *Subject Test	RFI*	11:28
04 Room Air	0.000	11:29

*Invalid Test
Inhibited - RFI

Sub Name = TEST, BOBBY NONE
Sub DOB = 01/01/1985
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/TES851012
Lot No = 02621080A1
Cyl No = 22
Expiration Date = 03/05/2023
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.14.00 09/16
05/20/2021 11:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:31
02 Std. Gas	0.080	11:31
03 Room Air	0.000	11:32
04 Std. Gas	0.080	11:32
05 Room Air	0.000	11:33
06 Std. Gas	0.080	11:33
07 Room Air	0.000	11:34

Lot No = 02621080A1
Cyl No = 22
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.14.00 09/16
05/20/2021 11:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:40
02 Std. Gas	0.080	11:41
03 Room Air	0.000	11:41
04 Std. Gas	0.080	11:41
05 Room Air	0.000	11:42
06 Std. Gas	0.080	11:42
07 Room Air	0.000	11:43

Lot No = 02621080A1
Cyl No = 22
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.14.00 09/16
05/20/2021 11:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:44
02 Std. Gas	0.080	11:45
03 Room Air	0.000	11:45
04 Std. Gas	0.080	11:46
05 Room Air	0.000	11:46
06 Std. Gas	0.080	11:47
07 Room Air	0.000	11:47

Lot No = 02621080A1
Cyl No = 22
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000