

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004947 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 728 psi Regulator: 725 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.030 AC
Sim. Ser #: MP5289 Lot #: 201911E Exp. Date: 11.19.21
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP5317 Lot #: 202012A Exp. Date: 12.8.22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: IC56 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 24119080A1 Cyl #: 31 Exp. Date: 11.5.21

Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.080</u>
Test 3 <u>0.079</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.079</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED
SIMULATOR RETURN O-RINGS.

Instrument is acceptable to be used in the field.

Chanda E. E...
Breath Analyst Signature

5/25/2021
Date

Kari Y. Heib
Reviewed by

5/26/21
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
05/25/2021 15:08

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05620080A1
Standard Cyl #? 11
Standard Expiration? 04/05/2022
Oper No? 666666

Flow Cal. Date: 04/07/2020
Slope 618
Intercept -565524

IR Calibration Date: 04/07/2020
 3um 9um

0th Coef(*100): -22371 -26151
1st Coef(*100): 260822 133081
2nd Coef(*100): 1936 1144
H2O adj(mg/l*10k): 495 535

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
05/25/2021 15:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:53
02 Std. Sol.	0.028	15:54
03 Room Air	0.000	15:55
04 Std. Sol.	0.028	15:55
05 Room Air	0.000	15:56
06 Std. Sol.	0.029	15:57
07 Room Air	0.000	15:57

08 Sim Temp = 34.0°C

Simul Ser No = MP5289
Std Sol No = 201911E
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.030 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
05/25/2021 16:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:20
02 Std. Sol.	0.297	16:21
03 Room Air	0.000	16:22
04 Std. Sol.	0.298	16:22
05 Room Air	0.000	16:23
06 Std. Sol.	0.297	16:24
07 Room Air	0.000	16:24

08 Sim Temp = 34.0°C

Simul Ser No = MP5317
Std Sol No = 202012A
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
05/25/2021 16:27

Test	AC	Time
01 Room Air	0.000	16:28
02 *Subject Test	INT*	16:28
03 Room Air	0.000	16:29

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 24119080A1
Cyl No = 31
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
05/25/2021 16:31

Test	AC	Time
01 Diagnostic	OK	16:32
02 Room Air	0.000	16:32
03 *Subject Test	RFI*	16:32
04 Room Air	0.000	16:33

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 24119080A1
Cyl No = 31
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
05/25/2021 16:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:34
02 Std. Gas	0.079	16:34
03 Room Air	0.000	16:34
04 Std. Gas	0.079	16:35
05 Room Air	0.000	16:35
06 Std. Gas	0.079	16:36
07 Room Air	0.000	16:36

Lot No = 24119080A1
Cyl No = 31
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
05/25/2021 16:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:37
02 Std. Gas	0.079	16:38
03 Room Air	0.000	16:38
04 Std. Gas	0.079	16:38
05 Room Air	0.000	16:39
06 Std. Gas	0.080	16:39
07 Room Air	0.000	16:40

Lot No = 24119080A1
Cyl No = 31
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
05/25/2021 16:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:41
02 Std. Gas	0.079	16:41
03 Room Air	0.000	16:42
04 Std. Gas	0.080	16:42
05 Room Air	0.000	16:43
06 Std. Gas	0.080	16:43
07 Room Air	0.000	16:44

Lot No = 24119080A1
Cyl No = 31
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC

Form 106-I8000