



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-004953</b>	Instrument Location <b>Pierce Co. Jail</b>
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input checked="" type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) <b>New to Loc 5-26-2021</b>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location: Level 2, Function E)
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): **None**
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <b>Darren Healdred</b>	Date <b>5/26/2021</b>
---	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <b>Charles E. Egan</b>	Certified Date <b>5-26-2021</b>
---	------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004953  
Location = RUGB                      8164.14.00 09/16  
05/26/2021    14:42

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            05620080A1  
Standard Cyl #?            043  
Standard Expiration?       04/05/2022  
Oper No?                    130392

Flow Cal. Date:             04/20/2018  
Slope                        609  
Intercept                   -462839

IR Calibration Date:        04/20/2018  
                                    3um                      9um

0th Coef(*100):	-22916	-27195
1st Coef(*100):	261768	134580
2nd Coef(*100):	1749	1216
H2O adj(mg/l*10k):	496	540

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DARREN HEIDBREDER

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = RUGB      8164.14.00 09/16  
05/26/2021      14:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:44
02 Std. Gas	0.080	14:45
03 Room Air	0.000	14:45
04 Std. Gas	0.080	14:45
05 Room Air	0.000	14:46
06 Std. Gas	0.081	14:46
07 Room Air	0.000	14:47

Lot No = 05620080A1  
Cyl No = 043  
Exp Date = 04/05/2022  
County = 35      Oper No. = 130392

  
Operator Signature  
DARREN HEIDBREDER

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = RUGB      8164.14.00 09/16  
05/26/2021      14:48

Test	AC	Time
01 Room Air	0.000	14:50
02 *Subject Test	RFI*	14:52
03 Room Air	0.000	14:52

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, TEST TEST  
Sub DOB = 09/05/1990  
Sub Sex = Male      Weight = 200  
Test = DUI      Cit = 1  
Dr. Lic. = ND/TES010001  
Lot No = 05620080A1  
Cyl No = 043  
Expiration Date = 04/05/2022  
County = 35      Oper No. = 130392



Operator Signature  
DARREN HEIDBREDER

Remarks:

Form 106-I8000