

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006495

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1000 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.030 AC
Sim. Ser #: MP5289 Lot #: 201911E Exp. Date: 11-19-21
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3062 Lot #: 202103E Exp. Date: 3-24-23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0-10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: ICS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 13 Exp. Date: 3-5-23

Test 1	<u>0.082</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.082</u>	Test 5	<u>0.082</u>	Test 8	<u>0.082</u>
Test 3	<u>0.082</u>	Test 6	<u>0.082</u>	Test 9	<u>0.082</u>
Average	<u>0.082</u>				

CEE

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edwards
Breath Analyst Signature

5/27/2021
Date

Kalid H. H. H.
Reviewed by

6/1/21
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006495
Location = TOXL 8164.14.00 09/16
05/27/2021 16:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:01
02 Std. Sol.	0.029	16:02
03 Room Air	0.000	16:03
04 Std. Sol.	0.030	16:03
05 Room Air	0.000	16:04
06 Std. Sol.	0.030	16:05
07 Room Air	0.000	16:05

08 Sim Temp = 34.0°C

Simul Ser No = MP5289
Std Sol No = 201911E
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.030 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006495
Location = TOXL 8164.14.00 09/16
05/27/2021 16:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:08
02 Std. Sol.	0.406	16:09
03 Room Air	0.000	16:09
04 Std. Sol.	0.407	16:10
05 Room Air	0.000	16:11
06 Std. Sol.	0.405	16:11
07 Room Air	0.000	16:12

08 Sim Temp = 34.0°C

Simul Ser No = MP3062
Std Sol No = 202103E
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.400 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006495
Location = TOXL 8164.14.00 09/16
05/27/2021 16:26

Test	AC	Time
01 Room Air	0.000	16:26
02 *Subject Test	INT*	16:27
03 Room Air	0.000	16:28

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 13
Expiration Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006495
Location = TOXL 8164.14.00 09/16
05/27/2021 16:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:31
02 Std. Gas	0.082	16:31
03 Room Air	0.000	16:32
04 Std. Gas	0.082	16:32
05 Room Air	0.000	16:33
06 Std. Gas	0.082	16:33
07 Room Air	0.000	16:34

Lot No = 02621080A1
Cyl No = 13
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006495
Location = TOXL 8164.14.00 09/16
05/27/2021 16:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:35
02 Std. Gas	0.081	16:35
03 Room Air	0.000	16:35
04 Std. Gas	0.082	16:36
05 Room Air	0.000	16:36
06 Std. Gas	0.082	16:37
07 Room Air	0.000	16:37

Lot No = 02621080A1
Cyl No = 13
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006495
Location = TOXL 8164.14.00 09/16
05/27/2021 16:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:38
02 Std. Gas	0.081	16:39
03 Room Air	0.000	16:39
04 Std. Gas	0.082	16:40
05 Room Air	0.000	16:40
06 Std. Gas	0.082	16:41
07 Room Air	0.000	16:41

Lot No = 02621080A1
Cyl No = 13
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000