



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-006501	Instrument Location Grafton PD
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) * ANNUAL INSPECTION *	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 27 MAY 2021
-------------------------------	----------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date 6.4.2021
---------------------------	-----------------------------------

Intoxilyzer Test Record and Checklist
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
 North Dakota Model 8000 SN 80-006501
 Location = GFTN 8164.14.00 09/16
 05/27/2021 18:09

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890- = |
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890- = |
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
 Start Test Sequence: DABACABA
 Display Prelim Rslt? Yes
 Display Third Digit? Yes
 Inhib Printer(Y/N)? No
 Display Volume? No
 Disable On Memfull? Yes
 # of Print Copies? 1
 Select Std (D/W/I)? Dry
 Standard Value? 0.080
 Standard Lot #? 02621080A1
 Standard Cyl #? 18
 Standard Expiration? 03/05/2023
 Oper No? 020305

Flow Cal. Date: 06/13/2019
 Slope 651
 Intercept -777821

IR Calibration Date: 06/13/2019

	3um	9um
0th Coef(*100):	-20531	-14906
1st Coef(*100):	270252	139749
2nd Coef(*100):	1917	1284
H2O adj(mg/l*10k):	393	321

***** Printer Test End *****



Operator Signature
 ANTHONY DUMAS

Remarks: Printer test after install

* Annual Inspection *

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = GFTN 8164.14.00 09/16
05/27/2021 18:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:11
02 Std. Gas	0.079	18:11
03 Room Air	0.000	18:12
04 Std. Gas	0.079	18:12
05 Room Air	0.000	18:13
06 Std. Gas	0.079	18:13
07 Room Air	0.000	18:13

Lot No = 02621080A1
Cyl No = 18
Exp Date = 03/05/2023
County = 50 Oper No. = 020305



Operator Signature
ANTHONY DUMAS

Remarks: ACA Test after install
* Annual Inspection *
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = GFTN 8164.14.00 09/16
05/27/2021 18:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	18:15
02 Room Air	RFI*	18:15

*Invalid Test
Inhibited - RFI

Lot No = 02621080A1
Cyl No = 18
Exp Date = 03/05/2023
County = 50 Oper No. = 020305



Operator Signature
ANTHONY DUMAS

Remarks: RFI TEST after ^Rannual inspection
Post install

Form 106-I8000