



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59281 (06/2018)

Serial Number <i>80-005362</i>	Instrument Location <i>McKenzie County 1201 12th ST SE Wadford City</i>
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): *Install*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>5-28-2021</i>
---	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>[Signature]</i>	Certified Date <i>5.28.2021</i>
---	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = MCSO 8164.14.00 09/16
05/28/2021 08:24

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05620080A1
Standard Cyl #? 28
Standard Expiration? 04/05/2022
Oper No? 130626

Flow Cal. Date: 05/22/2020
Slope 659
Intercept -620990

IR Calibration Date: 05/26/2020
 3um 9um

	3um	9um
0th Coef(*100):	-20883	-25370
1st Coef(*100):	265508	144960
2nd Coef(*100):	936	714
H2O adj(mg/l*10k):	391	379


***** Printer Test End *****



Operator Signature
BRYAN NAGEL

STATE OF NORTH DAKOTA }
COUNTY OF MCKENZIE }^{SS}

I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Watford City, North Dakota.

on this 26 day of May 2021

McKenzie County Sheriff's Department

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = MCSO 8164.14.00 09/16
05/28/2021 08:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:53
02 Std. Gas	0.079	08:53
03 Room Air	0.000	08:53
04 Std. Gas	0.079	08:54
05 Room Air	0.000	08:54
06 Std. Gas	0.080	08:55
07 Room Air	0.000	08:55

Lot No = 05620080A1
Cyl No = 28
Exp Date = 04/05/2022
County = 27 Oper No. = 130626


Operator Signature
BRYAN NAGEL


Remarks:

Form 106-I8000

STATE OF NORTH DAKOTA }
COUNTY OF MCKENZIE }^{ss}

I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Watford City, North Dakota.

on this 28 day of May 2021


McKenzie County Sheriff's Department


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = MCSO 8164.14.00 09/16
05/28/2021 08:57

Test	AC	Time
01 Room Air	RFI*	08:58
02 Room Air	0.000	08:59

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 04/11/1988
Sub Sex = Male Weight = 123
Test = DUI Cit = 123456789
Dr. Lic. = ND/TES881234
Lot No = 05620080A1
Cyl No = 28
Expiration Date = 04/05/2022
County = 27 Oper No. = 130626


Operator Signature
BRYAN NAGEL


Remarks:

RFI Test

Form 106-I8000

STATE OF NORTH DAKOTA }
COUNTY OF MCKENZIE }^{ss}

I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Watford City, North Dakota.

on this 28 day of May 2021

McKenzie County Sheriff's Department