

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004952

Location: Tox L

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 959 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3064 Lot #: 20070 Exp. Date: 2.13.22
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP5317 Lot #: 2020/2A Exp. Date: 12.8.22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: JCS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621 080A / Cyl #: 26 Exp. Date: 3.5.23
Test 1 0.079 Test 4 0.079 Test 7 0.079
Test 2 0.079 Test 5 0.079 Test 8 0.079
Test 3 0.079 Test 6 0.079 Test 9 0.080
Average 0.079

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED
SIMULATOR RETURN O-RING.

Instrument is acceptable to be used in the field.

Charles E. Ehr
Breath Analyst Signature

6/2/21
Date

Kali G. Heib
Reviewed by

6-7-21
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/02/2021 09:39

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 4
Standard Expiration? 11/05/2021
Oper No? 666666

Flow Cal. Date: 01/19/2010
Slope 694
Intercept -638616

IR Calibration Date: 05/18/2016

	3um	9um
0th Coef(*100):	-17808	-20871
1st Coef(*100):	256729	129583
2nd Coef(*100):	2826	1310
H2O adj(mg/l*10k):	344	422

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/02/2021 09:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:43
02 Std. Sol.	0.018	09:44
03 Room Air	0.000	09:44
04 Std. Sol.	0.018	09:45
05 Room Air	0.000	09:45
06 Std. Sol.	0.018	09:46
07 Room Air	0.000	09:47

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 20070

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/02/2021 09:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:51
02 Std. Sol.	0.295	09:52
03 Room Air	0.000	09:53
04 Std. Sol.	0.297	09:53
05 Room Air	0.000	09:54
06 Std. Sol.	0.298	09:55
07 Room Air	0.000	09:55

08 Sim Temp = 34.0°C

Simul Ser No = MP5317
Std Sol No = 202012A
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/02/2021 09:57

Test	AC	Time
01 Room Air	0.000	09:57
02 *Subject Test	INT*	09:58
03 Room Air	0.000	09:58

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 26
Expiration Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/02/2021 09:58

Test	AC	Time
01 Diagnostic	OK	09:59
02 Room Air	0.000	10:00
03 *Subject Test	RFI*	10:00
04 Room Air	0.000	10:00

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 26
Expiration Date = 03/05/2023
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/02/2021 10:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:01
02 Std. Gas	0.079	10:02
03 Room Air	0.000	10:02
04 Std. Gas	0.079	10:02
05 Room Air	0.000	10:03
06 Std. Gas	0.079	10:03
07 Room Air	0.000	10:04

Lot No = 02621080A1
Cyl No = 26
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/02/2021 10:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:05
02 Std. Gas	0.079	10:05
03 Room Air	0.000	10:05
04 Std. Gas	0.079	10:06
05 Room Air	0.000	10:06
06 Std. Gas	0.079	10:07
07 Room Air	0.000	10:07

Lot No = 02621080A1
Cyl No = 26
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/02/2021 10:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:08
02 Std. Gas	0.079	10:08
03 Room Air	0.000	10:09
04 Std. Gas	0.079	10:09
05 Room Air	0.000	10:10
06 Std. Gas	0.080	10:10
07 Room Air	0.000	10:11

Lot No = 02621080A1
Cyl No = 26
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000