

**INTOXILYZER® 8000 INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-004191 Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 193 psi Regulator: 225 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP3064 Lot #: 20070 Exp. Date: 2.13.22
    - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP5317 Lot #: 202012A Exp. Date: 12.8.22

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3847 Lot #: ICS 6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 24119080A1 Cyl #: 31 Exp. Date: 11.5.21

Test 1	<u>0.080</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.081</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.081</u>	Test 9	<u>0.081</u>
Average	<u>0.081</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

---

---

---

Instrument is acceptable to be used in the field.

Charles E. Eck  
Breath Analyst Signature

6/2/21  
Date

Kali L. Hib  
Reviewed by

6-7-21  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004191  
Location = TOXL                8164.14.00 09/16  
06/02/2021                      15:47

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            05620080A1  
Standard Cyl #?            6  
Standard Expiration?       04/05/2022  
Oper No?                    666666

Flow Cal. Date:            01/28/2010  
Slope                      738  
Intercept                  -777574

IR Calibration Date:       04/20/2017  
                                    3um                      9um

-----  
0th Coef(\*100):            -8758                      -19235  
1st Coef(\*100):            262182                     129903  
2nd Coef(\*100):            3907                        1462  
H2O adj(mg/l\*10k):        423                         588

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = TOXL      8164.14.00 09/16  
06/02/2021      15:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:50
02 Std. Sol.	0.020	15:50
03 Room Air	0.000	15:51
04 Std. Sol.	0.020	15:52
05 Room Air	0.000	15:52
06 Std. Sol.	0.020	15:53
07 Room Air	0.000	15:53

08 Sim Temp = 34.0°C

Simul Ser No = MP3064  
Std Sol No = 20070  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-18000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004191  
Location = TOXL                 8164.14.00 09/16  
06/02/2021                        16:14

Test	AC	Time
01 Room Air	0.000	16:14
02 *Subject Test	INT*	16:15
03 Room Air	0.000	16:15

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                     Weight = NA  
Test = DUI                         Cit = INTERFERENT  
Dr. Lic. = ND/TES989643  
Lot No = 24119080A1  
Cyl No = 31  
Expiration Date = 11/05/2021  
County = 08                        Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004191  
Location = TOXL                8164.14.00 09/16  
06/02/2021                        16:16

Test	AC	Time
01 Diagnostic	OK	16:16
02 Room Air	0.000	16:17
03 *Subject Test	RFI*	16:17
04 Room Air	0.000	16:17

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                    Weight = NA  
Test = DUI                        Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 24119080A1  
Cyl No = 31  
Expiration Date = 11/05/2021  
County = 08                        Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = TOXL      8164.14.00 09/16  
06/02/2021      16:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:19
02 Std. Gas	0.080	16:19
03 Room Air	0.000	16:19
04 Std. Gas	0.081	16:20
05 Room Air	0.000	16:20
06 Std. Gas	0.080	16:21
07 Room Air	0.000	16:21

Lot No = 24119080A1  
Cyl No = 31  
Exp Date = 11/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC  
Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = TOXL      8164.14.00 09/16  
06/02/2021      16:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:22
02 Std. Gas	0.081	16:22
03 Room Air	0.000	16:23
04 Std. Gas	0.081	16:23
05 Room Air	0.000	16:24
06 Std. Gas	0.081	16:24
07 Room Air	0.000	16:25

Lot No = 24119080A1  
Cyl No = 31  
Exp Date = 11/05/2021  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000



